

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032817\
 Data File : BE092859.D
 Acq On : 29 Mar 2017 5:29
 Operator : SJ/MA
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Mar 29 06:41:50 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 28 14:58:05 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.77	152	7627	5.00	ng	-0.02
7) Naphthalene-d8	10.54	136	34595	5.00	ng	0.00
12) Acenaphthene-d10	14.40	164	17392	5.00	ng	0.00
18) Phenanthrene-d10	17.13	188	38840	5.00	ng	-0.02
24) Chrysene-d12	21.28	240	46287	5.00	ng	-0.02
31) Perylene-d12	23.70	264	38701	5.00	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.36	112	698	0.39	ng	0.00
5) Phenol-d6	6.94	99	1500	0.58	ng	0.00
8) Nitrobenzene-d5	8.93	82	826	0.42	ng	0.00
13) 2,4,6-Tribromophenol	15.88	330	104	0.38	ng	0.00
14) 2-Fluorobiphenyl	13.02	172	2318	0.40	ng	-0.01
26) Terphenyl-d14	19.74	244	2745	0.34	ng	-0.02
Target Compounds						Qvalue
2) 1,4-Dioxane	3.31	88	433	0.42	ng	93
3) n-Nitrosodimethylamine	3.61	42	431	0.41	ng	# 85
6) bis(2-Chloroethyl)ether	7.20	93	1025	0.41	ng	# 50
9) Naphthalene	10.58	128	3064	0.40	ng	98
10) Hexachlorobutadiene	10.89	225	405	0.40	ng	99
11) 2-Methylnaphthalene	12.21	142	1887	0.39	ng	90
15) Acenaphthylene	14.09	152	9575	0.36	ng	100
16) Acenaphthene	14.45	154	1817	0.38	ng	85
17) Fluorene	15.44	166	2140	0.38	ng	97
19) Hexachlorobenzene	16.47	284	622	0.43	ng	# 65
20) Pentachlorophenol	16.81	266	137	0.51	ng	91
21) Phenanthrene	17.18	178	4399	0.47	ng	# 95
22) Anthracene	17.27	178	3421	0.40	ng	100
23) Fluoranthene	19.18	202	19954	0.46	ng	# 74
25) Pyrene	19.54	202	16440	0.36	ng	99
27) Benzo(a)anthracene	21.26	228	3756	0.34	ng	# 95
28) Chrysene	21.32	228	4735	0.41	ng	# 77
29) Bis(2-ethylhexyl)phthalate	21.21	149	983	0.37	ng	# 76
30) Indeno(1,2,3-cd)pyrene	26.24	276	14205	0.33	ng	# 93
32) Benzo(b)fluoranthene	22.96	252	3557	0.35	ng	98
33) Benzo(k)fluoranthene	23.01	252	3372	0.37	ng	# 95
34) Benzo(a)pyrene	23.59	252	2822	0.33	ng	# 96
35) Dibenzo(a,h)anthracene	26.27	278	3093	0.34	ng	# 96
36) Benzo(g,h,i)perylene	27.02	276	12977	0.35	ng	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032817\
Data File : BE092859.D
Acq On : 29 Mar 2017 5:29
Operator : SJ/MA
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampled :

Quant Time: Mar 29 06:41:50 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032817.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 28 14:58:05 2017
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 5.00 ng
 RT: 7.77 min Scan# 555
 Delta R.T. -0.02 min
 Lab File: BE092859.D
 Acq: 29 Mar 2017 5:29




Instrument :
 BNA_E
 ClientSampled :


Tot Ion:152 Resp: 7627
 Ion Ratio Lower Upper
 152 100
 150 184.4 106.0 159.0#
 115 94.9 31.2 46.8#



#2
 1,4-Dioxane
 Concen: 0.42 ng
 RT: 3.31 min Scan# 62
 Delta R.T. -0.01 min
 Lab File: BE092859.D
 Acq: 29 Mar 2017 5:29

Tgt Ion: 88 Resp: 433
 Ion Ratio Lower Upper
 88 100
 58 74.4 63.6 95.4
 43 28.6 28.0 42.0









#3
n-Nitrosodimethylamine
Concen: 0.41 ng
RT: 3.61 min Scan# 88
Delta R.T. 0.00 min
Lab File: BE092859.D
Acq: 29 Mar 2017 5:29




Instrument :
BNA_E
ClientSampleId :


Tot Ion: 42 Resp: 431
Ion Ratio Lower Upper
42 100
74 123.9 86.0 129.0
44 10.0 5.1 7.7#



#4
2-Fluorophenol
Concen: 0.39 ng
RT: 5.36 min Scan# 267
Delta R.T. -0.00 min
Lab File: BE092859.D
Acq: 29 Mar 2017 5:29

Tgt Ion: 112 Resp: 698
Ion Ratio Lower Upper
112 100
64 69.2 53.5 80.3
63 35.5 27.9 41.9





#5

Phenol-d6

Concen: 0.58 ng

RT: 6.94 min Scan# 519

Delta R.T. -0.00 min

Lab File: BE092859.D

Acq: 29 Mar 2017 5:29

Instrument :

BNA_E

ClientSampleId :


Tot Ion: 99 Resp: 1500

Ion Ratio Lower Upper

99 100

42 0.0 16.0 24.0#




71 0.0 30.6 45.8#



Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0



#6

bis(2-Chloroethyl)ether

Concen: 0.41 ng

RT: 7.20 min Scan# 530

Delta R.T. -0.00 min

Lab File: BE092859.D

Acq: 29 Mar 2017 5:29


Tgt Ion: 93 Resp: 1025

Ion Ratio Lower Upper

93 100

63 77.4 16.0 24.0#



95 30.3 24.0 36.0




Abundance Ion 93.00 (92.70 to 93.70): BE0

Ion 63.00 (62.70 to 63.70): BE0

Ion 95.00 (94.70 to 95.70): BE0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.54 min Scan# 679

Delta R.T. -0.00 min

Lab File: BE092859.D

Acq: 29 Mar 2017 5:29

Instrument :

BNA_E

ClientSampleId :

Tot Ion:136 Resp: 34595


Ion Ratio Lower Upper

136 100

137 12.0 8.2 12.4

54 4.4 9.6 14.4#

68 4.5 6.7 10.1#






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0



#8

Nitrobenzene-d5

Concen: 0.42 ng

RT: 8.93 min Scan# 605

Delta R.T. -0.00 min

Lab File: BE092859.D

Acq: 29 Mar 2017 5:29


Tgt Ion: 82 Resp: 826

Ion Ratio Lower Upper

82 100

128 61.0 39.0 58.4#



54 37.1 44.8 67.2#




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0





#9

Naphthalene

Concen: 0.40 ng

RT: 10.58 min Scan# 681

Delta R.T. -0.00 min

Lab File: BE092859.D

Acq: 29 Mar 2017 5:29

Instrument :

BNA_E

ClientSampleId :





Tot Ion:128 Resp: 3064

Ion Ratio Lower Upper

128 100

129 12.5 11.2 16.8

127 15.0 11.6 17.4



#10

Hexachlorobutadiene

Concen: 0.40 ng

RT: 10.89 min Scan# 697

Delta R.T. -0.00 min

Lab File: BE092859.D

Acq: 29 Mar 2017 5:29




Tgt Ion:225 Resp: 405


Ion Ratio Lower Upper

225 100

223 63.0 50.6 76.0

227 63.2 51.4 77.0









#11
2-Methylnaphthalene
Concen: 0.39 ng
RT: 12.21 min Scan# 770
Delta R.T. 0.00 min
Lab File: BE092859.D
Acq: 29 Mar 2017 5:29




Instrument :
BNA_E
ClientSampleId :


Tot Ion:142 Resp: 1887
Ion Ratio Lower Upper
142 100
141 84.4 73.7 110.5
115 34.3 34.0 51.0



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.40 min Scan# 933
Delta R.T. 0.00 min
Lab File: BE092859.D
Acq: 29 Mar 2017 5:29

Tgt Ion:164 Resp: 17392
Ion Ratio Lower Upper
164 100
162 91.5 71.4 107.0
160 37.5 27.4 41.2









#13
 2,4,6-Tribromophenol
 Concen: 0.38 ng
 RT: 15.88 min Scan# 1042
 Delta R.T. -0.00 min
 Lab File: BE092859.D
 Acq: 29 Mar 2017 5:29

Instrument :
 BNA_E
 ClientSampled :

Tot Ion:330 Resp: 104
 Ion Ratio Lower Upper
 330 100
 332 95.2 75.4 113.0
 141 59.6 31.0 46.6#




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#14
 2-Fluorobiphenyl
 Concen: 0.40 ng
 RT: 13.02 min Scan# 840
 Delta R.T. -0.01 min
 Lab File: BE092859.D
 Acq: 29 Mar 2017 5:29

Tgt Ion:172 Resp: 2318
 Ion Ratio Lower Upper
 172 100
 171 35.7 30.1 45.1
 170 23.8 21.8 32.6



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#15
Acenaphthylene
Concen: 0.36 ng
RT: 14.09 min Scan# 915
Delta R.T. -0.02 min
Lab File: BE092859.D
Acq: 29 Mar 2017 5:29




Instrument :
BNA_E
ClientSampleId :


Tot Ion:152 Resp: 9575
Ion Ratio Lower Upper
152 100
151 4.9 3.9 5.9
153 13.1 10.4 15.6



#16
Acenaphthene
Concen: 0.38 ng
RT: 14.45 min Scan# 936
Delta R.T. 0.00 min
Lab File: BE092859.D
Acq: 29 Mar 2017 5:29

Tgt Ion:154 Resp: 1817
Ion Ratio Lower Upper
154 100
153 477.8 350.8 526.2
152 236.1 171.1 256.7









#17
Fluorene
 Concen: 0.38 ng
 RT: 15.44 min Scan# 1004
 Delta R.T. -0.00 min
 Lab File: BE092859.D
 Acq: 29 Mar 2017 5:29




Instrument :
 BNA_E
 ClientSampleId :


Tot Ion:166 Resp: 2140
 Ion Ratio Lower Upper
 166 100
 165 101.5 78.2 117.4
 167 14.1 11.0 16.4



#18
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 17.13 min Scan# 1109
 Delta R.T. -0.02 min
 Lab File: BE092859.D
 Acq: 29 Mar 2017 5:29

Tgt Ion:188 Resp: 38840
 Ion Ratio Lower Upper
 188 100
 94 21.7 3.2 4.8#
 80 23.8 2.6 3.8#









#19
Hexachlorobenzene
Concen: 0.43 ng
RT: 16.47 min Scan# 1080
Delta R.T. -0.00 min
Lab File: BE092859.D
Acq: 29 Mar 2017 5:29




Instrument :
BNA_E
ClientSampled :


Tot Ion:284 Resp: 622
Ion Ratio Lower Upper
284 100
142 0.0 29.1 43.7#
249 31.2 27.9 41.9



#20
Pentachlorophenol
Concen: 0.51 ng
RT: 16.81 min Scan# 1095
Delta R.T. 0.00 min
Lab File: BE092859.D
Acq: 29 Mar 2017 5:29

Tgt Ion:266 Resp: 137
Ion Ratio Lower Upper
266 100
268 91.0 62.5 93.7
264 73.0 57.2 85.8





#21

Phenanthrene

Concen: 0.47 ng

RT: 17.18 min Scan# 1111

Delta R.T. -0.00 min





Lab File: BE092859.D

Acq: 29 Mar 2017 5:29

Instrument :
BNA_E
ClientSampleId :

Tot Ion:178 Resp: 4399

Ion	Ratio	Lower	Upper
178	100		
176	19.3	15.8	23.8
179	15.5	16.0	24.0



#22

Anthracene

Concen: 0.40 ng

RT: 17.27 min Scan# 1115




Delta R.T. -0.00 min


Lab File: BE092859.D

Acq: 29 Mar 2017 5:29

Tgt Ion:178 Resp: 3421

Ion	Ratio	Lower	Upper
178	100		
176	18.4	15.0	22.4
179	15.1	12.2	18.2









#23
 Fluoranthene
 Concen: 0.46 ng
 RT: 19.18 min Scan# 1199
 Delta R.T. 0.00 min
 Lab File: BE092859.D
 Acq: 29 Mar 2017 5:29




Instrument :
 BNA_E
 ClientSampleId :


Tot Ion:202 Resp: 19954
 Ion Ratio Lower Upper
 202 100
 101 6.3 2.2 3.4#
 203 29.5 13.7 20.5#



#24
 Chrysene-d12
 Concen: 5.00 ng
 RT: 21.28 min Scan# 1320
 Delta R.T. -0.02 min
 Lab File: BE092859.D
 Acq: 29 Mar 2017 5:29

Tgt Ion:240 Resp: 46287
 Ion Ratio Lower Upper
 240 100
 120 6.4 2.6 4.0#
 236 28.3 24.6 37.0





#25

Pvrene

Concen: 0.36 ng

RT: 19.54 min Scan# 1220

Delta R.T. 0.00 min

Lab File: BE092859.D

Acq: 29 Mar 2017 5:29

Instrument :

BNA_E

ClientSampleId :


Tot Ion:202 Resp: 16440

Ion Ratio Lower Upper

202 100

200 5.0 4.2 6.4




203 17.8 13.9 20.9



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#26

Terphenyl-d14

Concen: 0.34 ng

RT: 19.74 min Scan# 1232

Delta R.T. -0.02 min

Lab File: BE092859.D

Acq: 29 Mar 2017 5:29


Tgt Ion:244 Resp: 2745

Ion Ratio Lower Upper

244 100

212 12.3 6.7 10.1#



122 18.5 3.6 5.4#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E









#27
Benzo(a)anthracene
Concen: 0.34 ng
RT: 21.26 min Scan# 1319
Delta R.T. 0.00 min
Lab File: BE092859.D
Acq: 29 Mar 2017 5:29




Instrument :
BNA_E
ClientSampled :


Tot Ion:228 Resp: 3756
Ion Ratio Lower Upper
228 100
226 27.7 23.6 35.4
229 20.3 13.3 19.9#



#28
Chrysene
Concen: 0.41 ng
RT: 21.32 min Scan# 1322
Delta R.T. 0.00 min
Lab File: BE092859.D
Acq: 29 Mar 2017 5:29

Tgt Ion:228 Resp: 4735
Ion Ratio Lower Upper
228 100
226 29.5 36.3 54.5#
229 20.2 10.6 16.0#









#29
 Bis(2-ethylhexyl)phthalate
 Concen: 0.37 ng
 RT: 21.21 min Scan# 1316
 Delta R.T. -0.02 min
 Lab File: BE092859.D
 Acq: 29 Mar 2017 5:29




Instrument :
 BNA_E
 ClientSampleId :


Tot Ion:149 Resp: 983
 Ion Ratio Lower Upper
 149 100
 167 24.5 16.0 24.0#
 279 2.6 16.0 24.0#



#30
 Indeno(1,2,3-cd)pyrene
 Concen: 0.33 ng
 RT: 26.24 min Scan# 1703
 Delta R.T. -0.02 min
 Lab File: BE092859.D
 Acq: 29 Mar 2017 5:29

Tgt Ion:276 Resp: 14205
 Ion Ratio Lower Upper
 276 100
 138 31.9 20.3 30.5#
 277 24.6 19.7 29.5









#31
 Pervlene-d12
 Concen: 5.00 ng
 RT: 23.70 min Scan# 1492
 Delta R.T. -0.01 min
 Lab File: BE092859.D
 Acq: 29 Mar 2017 5:29




Instrument :
 BNA_E
 ClientSampleId :


Tot Ion:264 Resp: 38701
 Ion Ratio Lower Upper
 264 100
 260 26.5 20.1 30.1
 265 20.5 17.9 26.9



#32
 Benzo(b)fluoranthene
 Concen: 0.35 ng
 RT: 22.96 min Scan# 1428
 Delta R.T. -0.01 min
 Lab File: BE092859.D
 Acq: 29 Mar 2017 5:29

Tgt Ion:252 Resp: 3557
 Ion Ratio Lower Upper
 252 100
 253 23.8 18.7 28.1
 125 14.8 10.5 15.7





#33

Benzo(k)fluoranthene

Concen: 0.37 ng

RT: 23.01 min Scan# 1432

Delta R.T. -0.01 min

Lab File: BE092859.D

Acq: 29 Mar 2017 5:29

Instrument :

BNA_E

ClientSampleId :


Tot Ion:252 Resp: 3372

Ion Ratio Lower Upper

252 100

253 23.6 20.3 30.5




125 15.0 9.6 14.4#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#34

Benzo(a)pyrene

Concen: 0.33 ng

RT: 23.59 min Scan# 1482

Delta R.T. -0.01 min

Lab File: BE092859.D

Acq: 29 Mar 2017 5:29


Tgt Ion:252 Resp: 2822

Ion Ratio Lower Upper

252 100

253 23.9 19.0 28.6



125 19.0 11.8 17.8#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#35
 Dibenzo(a,h)anthracene
 Concen: 0.34 ng
 RT: 26.27 min Scan# 1705
 Delta R.T. -0.02 min
 Lab File: BE092859.D
 Acq: 29 Mar 2017 5:29

Instrument :
 BNA_E
 ClientSampleId :

Tot Ion:278 Resp: 3093
 Ion Ratio Lower Upper
 278 100
 139 21.6 13.8 20.8#
 279 26.9 21.4 32.2



#36
 Benzo(g,h,i)perylene
 Concen: 0.35 ng
 RT: 27.02 min Scan# 1749
 Delta R.T. -0.00 min
 Lab File: BE092859.D
 Acq: 29 Mar 2017 5:29

Tgt Ion:276 Resp: 12977
 Ion Ratio Lower Upper
 276 100
 277 28.0 19.8 29.8
 138 28.0 19.8 29.6

