

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032817\
 Data File : BE092836.D
 Acq On : 28 Mar 2017 14:25
 Operator : SJ/MA
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE032817

Quant Time: Mar 28 15:01:45 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 28 14:58:05 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.668	0.709	-6.1	101	-0.01
3	n-Nitrosodimethylamine	0.691	0.662	4.2	101	0.00
4 S	2-Fluorophenol	1.164	1.121	3.7	96	0.00
5 S	Phenol-d6	1.708	1.620	5.2	100	0.00
6	bis(2-Chloroethyl)ether	1.625	1.627	-0.1	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.287	0.279	2.8	109	0.00
9	Naphthalene	1.095	1.092	0.3	99	0.00
10	Hexachlorobutadiene	0.147	0.144	2.0	99	0.00
11	2-Methylnaphthalene	0.695	0.673	3.2	99	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
13 S	2,4,6-Tribromophenol	0.078	0.064	17.9	90	0.00
14 S	2-Fluorobiphenyl	1.648	1.666	-1.1	99	-0.01
15	Acenaphthylene	7.666	6.703	12.6	97	0.00
16	Acenaphthene	1.367	1.308	4.3	98	0.00
17	Fluorene	1.629	1.558	4.4	96	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
19	Hexachlorobenzene	0.185	0.192	-3.8	101	0.00
20	Pentachlorophenol	0.048	0.036	25.0#	83	0.00
21	Phenanthrene	1.217	1.331	-9.4	101	0.00
22	Anthracene	1.106	1.078	2.5	100	0.00
23	Fluoranthene	5.597	5.122	8.5	94	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	95	-0.02
25	Pyrene	4.972	4.631	6.9	94	0.00
26 S	Terphenyl-d14	0.864	0.805	6.8	92	0.00
27	Benzo(a)anthracene	1.185	1.105	6.8	97	0.00
28	Chrysene	1.250	1.230	1.6	95	0.00
29	Bis(2-ethylhexyl)phthalate	0.362	0.262	27.6#	87	0.00
30	Indeno(1,2,3-cd)pyrene	4.702	4.296	8.6	93	0.00
31 I	Perylene-d12	1.000	1.000	0.0	96	0.00
32	Benzo(b)fluoranthene	1.331	1.138	14.5	85	-0.01
33	Benzo(k)fluoranthene	1.192	1.176	1.3	100	0.00
34 C	Benzo(a)pyrene	1.100	0.956	13.1	90	0.00
35	Dibenzo(a,h)anthracene	1.170	1.065	9.0	92	0.00
36	Benzo(g,h,i)perylene	4.851	4.538	6.5	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0