

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032817\
 Data File : BE092845.D
 Acq On : 28 Mar 2017 20:56
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 29 05:58:01 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 28 14:58:05 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	-0.02
2	1,4-Dioxane	0.668	0.666	0.3	99	0.00
3	n-Nitrosodimethylamine	0.691	0.642	7.1	102	0.00
4 S	2-Fluorophenol	1.164	1.088	6.5	97	0.00
5 S	Phenol-d6	1.708	1.548	9.4	99	0.00
6	bis(2-Chloroethyl)ether	1.625	1.650	-1.5	105	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.287	0.277	3.5	111	0.00
9	Naphthalene	1.095	1.085	0.9	102	0.00
10	Hexachlorobutadiene	0.147	0.142	3.4	101	0.00
11	2-Methylnaphthalene	0.695	0.675	2.9	102	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
13 S	2,4,6-Tribromophenol	0.078	0.066	15.4	97	0.00
14 S	2-Fluorobiphenyl	1.648	1.685	-2.2	105	-0.01
15	Acenaphthylene	7.666	6.526	14.9	99	0.00
16	Acenaphthene	1.367	1.322	3.3	104	0.00
17	Fluorene	1.629	1.574	3.4	101	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	99	-0.02
19	Hexachlorobenzene	0.185	0.197	-6.5	107	0.00
20	Pentachlorophenol	0.048	0.040	16.7	96	0.00
21	Phenanthrene	1.217	1.438	-18.2	113	0.00
22	Anthracene	1.106	1.116	-0.9	108	0.00
23	Fluoranthene	5.597	6.692	-19.6	127	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	100	-0.02
25	Pyrene	4.972	4.614	7.2	100	0.00
26 S	Terphenyl-d14	0.864	0.792	8.3	96	0.00
27	Benzo(a)anthracene	1.185	1.098	7.3	102	0.00
28	Chrysene	1.250	1.252	-0.2	102	0.00
29	Bis(2-ethylhexyl)phthalate	0.362	0.238	34.3#	84	-0.02
30	Indeno(1,2,3-cd)pyrene	4.702	4.026	14.4	92	-0.02
31 I	Perylene-d12	1.000	1.000	0.0	100	0.00
32	Benzo(b)fluoranthene	1.331	1.160	12.8	91	-0.01
33	Benzo(k)fluoranthene	1.192	1.136	4.7	102	-0.01
34 C	Benzo(a)pyrene	1.100	0.903	17.9	89	0.00
35	Dibenzo(a,h)anthracene	1.170	1.041	11.0	95	-0.02
36	Benzo(g,h,i)perylene	4.851	4.423	8.8	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0