

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032817\
 Data File : BE092845.D
 Acq On : 28 Mar 2017 20:56
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 29 05:58:01 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 28 14:58:05 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	102	-0.02
2	1,4-Dioxane	0.400	0.399	0.3	99	0.00
3	n-Nitrosodimethylamine	0.400	0.372	7.0	102	0.00
4 S	2-Fluorophenol	0.400	0.374	6.5	97	0.00
5 S	Phenol-d6	0.400	0.362	9.5	99	0.00
6	bis(2-Chloroethyl)ether	0.400	0.406	-1.5	105	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.400	0.386	3.5	111	0.00
9	Naphthalene	0.400	0.396	1.0	102	0.00
10	Hexachlorobutadiene	0.400	0.386	3.5	101	0.00
11	2-Methylnaphthalene	0.400	0.388	3.0	102	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	103	0.00
13 S	2,4,6-Tribromophenol	0.400	0.336	16.0	97	0.00
14 S	2-Fluorobiphenyl	0.400	0.409	-2.2	105	-0.01
15	Acenaphthylene	0.400	0.341	14.8	99	0.00
16	Acenaphthene	0.400	0.387	3.3	104	0.00
17	Fluorene	0.400	0.386	3.5	101	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	99	-0.02
19	Hexachlorobenzene	0.400	0.424	-6.0	107	0.00
20	Pentachlorophenol	0.400	0.483	-20.7	96	0.00
21	Phenanthrene	0.400	0.473	-18.2	113	0.00
22	Anthracene	0.400	0.404	-1.0	108	0.00
23	Fluoranthene	0.400	0.478	-19.5	127	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	100	-0.02
25	Pyrene	0.400	0.371	7.3	100	0.00
26 S	Terphenyl-d14	0.400	0.367	8.3	96	0.00
27	Benzo(a)anthracene	0.400	0.371	7.3	102	0.00
28	Chrysene	0.400	0.401	-0.3	102	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.327	18.3	84	-0.02
30	Indeno(1,2,3-cd)pyrene	0.400	0.342	14.5	92	-0.02
31 I	Perylene-d12	5.000	5.000	0.0	100	0.00
32	Benzo(b)fluoranthene	0.400	0.349	12.8	91	-0.01
33	Benzo(k)fluoranthene	0.400	0.381	4.8	102	-0.01
34 C	Benzo(a)pyrene	0.400	0.328	18.0	89	0.00
35	Dibenzo(a,h)anthracene	0.400	0.356	11.0	95	-0.02
36	Benzo(g,h,i)perylene	0.400	0.365	8.8	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0