

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032817\
 Data File : BE092859.D
 Acq On : 29 Mar 2017 5:29
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 29 12:50:26 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 28 14:58:05 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	-0.02
2	1,4-Dioxane	0.668	0.710	-6.3	106	-0.01
3	n-Nitrosodimethylamine	0.691	0.706	-2.2	113	0.00
4 S	2-Fluorophenol	1.164	1.144	1.7	103	0.00
5 S	Phenol-d6	1.708	1.595	6.6	103	0.00
6	bis(2-Chloroethyl)ether	1.625	1.680	-3.4	108	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.287	0.298	-3.8	121	0.00
9	Naphthalene	1.095	1.107	-1.1	105	0.00
10	Hexachlorobutadiene	0.147	0.146	0.7	105	0.00
11	2-Methylnaphthalene	0.695	0.682	1.9	105	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
13 S	2,4,6-Tribromophenol	0.078	0.075	3.8	114	0.00
14 S	2-Fluorobiphenyl	1.648	1.666	-1.1	107	-0.01
15	Acenaphthylene	7.666	6.882	10.2	108	-0.02
16	Acenaphthene	1.367	1.306	4.5	106	0.00
17	Fluorene	1.629	1.538	5.6	102	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	104	-0.02
19	Hexachlorobenzene	0.185	0.200	-8.1	115	0.00
20	Pentachlorophenol	0.048	0.044	8.3	112	0.00
21	Phenanthrene	1.217	1.416	-16.4	117	0.00
22	Anthracene	1.106	1.101	0.5	112	0.00
23	Fluoranthene	5.597	6.422	-14.7	129	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	104	-0.02
25	Pyrene	4.972	4.440	10.7	99	0.00
26 S	Terphenyl-d14	0.864	0.741	14.2	93	-0.02
27	Benzo(a)anthracene	1.185	1.014	14.4	97	0.00
28	Chrysene	1.250	1.279	-2.3	108	0.00
29	Bis(2-ethylhexyl)phthalate	0.362	0.265	26.8#	96	-0.02
30	Indeno(1,2,3-cd)pyrene	4.702	3.836	18.4	91	-0.02
31 I	Perylene-d12	1.000	1.000	0.0	100	-0.01
32	Benzo(b)fluoranthene	1.331	1.149	13.7	89	-0.01
33	Benzo(k)fluoranthene	1.192	1.089	8.6	97	-0.01
34 C	Benzo(a)pyrene	1.100	0.911	17.2	90	-0.01
35	Dibenzo(a,h)anthracene	1.170	0.999	14.6	91	-0.02
36	Benzo(g,h,i)perylene	4.851	4.191	13.6	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0