

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032817\
 Data File : BE092859.D
 Acq On : 29 Mar 2017 5:29
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 29 12:50:26 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 28 14:58:05 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	103	-0.02
2	1,4-Dioxane	0.400	0.425	-6.2	106	-0.01
3	n-Nitrosodimethylamine	0.400	0.409	-2.2	113	0.00
4 S	2-Fluorophenol	0.400	0.393	1.8	103	0.00
5 S	Phenol-d6	0.400	0.373	6.8	103	0.00
6	bis(2-Chloroethyl)ether	0.400	0.413	-3.2	108	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.400	0.416	-4.0	121	0.00
9	Naphthalene	0.400	0.404	-1.0	105	0.00
10	Hexachlorobutadiene	0.400	0.399	0.3	105	0.00
11	2-Methylnaphthalene	0.400	0.392	2.0	105	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	107	0.00
13 S	2,4,6-Tribromophenol	0.400	0.383	4.3	114	0.00
14 S	2-Fluorobiphenyl	0.400	0.404	-1.0	107	-0.01
15	Acenaphthylene	0.400	0.359	10.3	108	-0.02
16	Acenaphthene	0.400	0.382	4.5	106	0.00
17	Fluorene	0.400	0.378	5.5	102	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	104	-0.02
19	Hexachlorobenzene	0.400	0.432	-8.0	115	0.00
20	Pentachlorophenol	0.400	0.511	-27.7#	112	0.00
21	Phenanthrene	0.400	0.466	-16.5	117	0.00
22	Anthracene	0.400	0.398	0.5	112	0.00
23	Fluoranthene	0.400	0.459	-14.8	129	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	104	-0.02
25	Pyrene	0.400	0.357	10.8	99	0.00
26 S	Terphenyl-d14	0.400	0.343	14.2	93	-0.02
27	Benzo(a)anthracene	0.400	0.343	14.2	97	0.00
28	Chrysene	0.400	0.409	-2.2	108	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.367	8.3	96	-0.02
30	Indeno(1,2,3-cd)pyrene	0.400	0.326	18.5	91	-0.02
31 I	Perylene-d12	5.000	5.000	0.0	100	-0.01
32	Benzo(b)fluoranthene	0.400	0.345	13.8	89	-0.01
33	Benzo(k)fluoranthene	0.400	0.366	8.5	97	-0.01
34 C	Benzo(a)pyrene	0.400	0.331	17.3	90	-0.01
35	Dibenzo(a,h)anthracene	0.400	0.342	14.5	91	-0.02
36	Benzo(g,h,i)perylene	0.400	0.346	13.5	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0