




Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
Data File : BE095901.D
Acq On : 29 Mar 2018 5:39
Operator : SJ/JU
Sample : J2060-01
Misc :
ALS Vial : 31 Sample Multiplier: 1

Quant Time: Mar 29 07:01:35 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 29 07:00:13 2018
Response via : Initial Calibration



TIC: BE095901.D

(16) Chrysene-d12 (I)

21.907min (+0.120) 0.40ng/ul

response 95

Ion	Exp%	Act%
240.00	100	100
120.00	20.60	124.64#
236.00	34.90	99.43#
0.00	0.00	0.00