




Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
Data File : BE095901.D
Acq On : 29 Mar 2018 5:39
Operator : SJ/JU
Sample : J2060-01
Misc :
ALS Vial : 31 Sample Multiplier: 1

Quant Time: Mar 29 07:50:14 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 29 07:00:13 2018
Response via : Initial Calibration



TIC: BE095901.D

(15) Fluoranthene (C)

19.706min (-19.706) 0.00ng/ul

response 0

Ion	Exp%	Act%
202.00	100	0.00
101.00	10.30	0.00
100.00	19.50	0.00
0.00	0.00	0.00