

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
 Data File : BE095879.D
 Acq On : 28 Mar 2018 16:34
 Operator : SJ/JU
 Sample : J1950-12
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Mar 29 04:40:17 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 28 16:31:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.41	152	1677	0.40	ng/ul	0.00
2) Naphthalene-d8	11.24	136	4572	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.00	164	2997	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.70	188	7385m	0.40	ng/ul	0.00
16) Chrysene-d12	21.79	240	7512m	0.40	ng/ul	0.00
20) Perylene-d12	24.42	264	7746m	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.79	152	2520	0.33	ng/ul	0.00
14) Fluoranthene-d10	19.68	212	8501m	0.36	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
Data File : BE095879.D
Acq On : 28 Mar 2018 16:34
Operator : SJ/JU
Sample : J1950-12
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Mar 29 04:40:17 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 28 16:31:37 2018
Response via : Initial Calibration

