

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
 Data File : BE095891.D
 Acq On : 28 Mar 2018 23:41
 Operator : SJ/JU
 Sample : J2059-08
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Mar 29 06:08:38 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 29 04:43:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.41	152	1704	0.40	ng/ul	0.00
2) Naphthalene-d8	11.24	136	5264	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.00	164	3322	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	7582m	0.40	ng/ul	0.00
16) Chrysene-d12	21.79	240	7818m	0.40	ng/ul	0.00
20) Perylene-d12	24.41	264	8468m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.79	152	3443	0.39	ng/ul	0.00
14) Fluoranthene-d10	19.68	212	11016m	0.46	ng/ul	0.00
Target Compounds						
8) Acenaphthene	15.05	153	662	0.052	ng/ul#	Ovalue 30

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
Data File : BE095891.D
Acq On : 28 Mar 2018 23:41
Operator : SJ/JU
Sample : J2059-08
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Mar 29 06:08:38 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 29 04:43:20 2018
Response via : Initial Calibration

