

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
 Data File : BE095900.D
 Acq On : 29 Mar 2018 5:03
 Operator : SJ/JU
 Sample : J1951-01
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Mar 29 07:38:59 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 29 07:00:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.42	152	1830	0.40	ng/ul	0.00
2) Naphthalene-d8	11.24	136	5608	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.00	164	3621	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.70	188	8075m	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	7883m	0.40	ng/ul	0.00
20) Perylene-d12	24.43	264	8088m	0.40	ng/ul	0.02
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.79	152	2056	0.22	ng/ul	0.00
14) Fluoranthene-d10	19.68	212	5351m	0.21	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	11.29	128	5014	0.320	ng/ul#	90
5) 2-Methylnaphthalene	12.87	142	1691	0.155	ng/ul	98
7) Acenaphthylene	14.73	152	783	0.043	ng/ul#	78
8) Acenaphthene	15.06	153	17159	1.246	ng/ul	95
9) Fluorene	16.03	166	10371	0.694	ng/ul	91
12) Phenanthrene	17.75	178	166739m	6.895	ng/ul	
13) Anthracene	17.84	178	43674m	1.931	ng/ul	
15) Fluoranthene	19.71	202	386954m	13.209	ng/ul	
17) Pyrene	20.07	202	385187	12.868	ng/ul#	78
18) Benzo(a)anthracene	21.77	228	285369m	11.219	ng/ul	
19) Chrysene	21.83	228	214551m	8.285	ng/ul	
21) Benzo(b)fluoranthene	23.61	252	309375m	11.043	ng/ul	
22) Benzo(k)fluoranthene	23.66	252	104703m	3.931	ng/ul	
23) Benzo(a)pyrene	24.31	252	227033m	8.876	ng/ul	
24) Indeno(1,2,3-cd)pyrene	27.24	276	155746	5.491	ng/ul#	72
25) Dibenzo(a,h)anthracene	27.24	278	40098m	1.768	ng/ul	
26) Benzo(a,h,i)perylene	28.11	276	151586	6.221	ng/ul#	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
Data File : BE095900.D
Acq On : 29 Mar 2018 5:03
Operator : SJ/JU
Sample : J1951-01
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Mar 29 07:38:59 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 29 07:00:13 2018
Response via : Initial Calibration

