

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
 Data File : BE095871.D
 Acq On : 28 Mar 2018 11:20
 Operator : SJ/JU
 Sample : PB107681BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SBLK81

Manual Integrations
 APPROVED

Sohil
 3/29/2018 5:17:27 PM

Quant Time: Mar 28 16:53:19 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 28 16:31:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.42	152	1689	0.40	ng/ul	0.00
2) Naphthalene-d8	11.24	136	4600	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.99	164	3067	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.70	188	7291m	0.40	ng/ul	0.00
16) Chrysene-d12	21.79	240	7976m	0.40	ng/ul	0.00
20) Perylene-d12	24.42	264	7126m	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.79	152	2773	0.36	ng/ul	0.00
14) Fluoranthene-d10	19.68	212	9414m	0.41	ng/ul	0.00

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
Data File : BE095871.D
Acq On : 28 Mar 2018 11:20
Operator : SJ/JU
Sample : PB107681BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SBLK81

Manual Integrations
APPROVED

Sohil
3/29/2018 5:17:27 PM

Quant Time: Mar 28 16:53:19 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 28 16:31:37 2018
Response via : Initial Calibration

