

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
 Data File : BE095873.D
 Acq On : 28 Mar 2018 12:59
 Operator : SJ/JU
 Sample : J2060-07
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C0AT3

Manual Integrations
 APPROVED

Sohil
 3/29/2018 5:17:34 PM

Quant Time: Mar 29 11:28:41 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 28 16:31:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.42	152	1650	0.40	ng/ul	0.00
2) Naphthalene-d8	11.24	136	4428	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.00	164	2945	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.70	188	7089m	0.40	ng/ul	0.00
16) Chrysene-d12	21.79	240	7643m	0.40	ng/ul	0.00
20) Perylene-d12	24.41	264	6943m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.80	152	2622	0.35	ng/ul	0.00
14) Fluoranthene-d10	19.68	212	8536m	0.38	ng/ul	0.00
Target Compounds				Ovalue		
12) Phenanthrene	17.75	178	1226m	0.058	ng/ul	
15) Fluoranthene	19.71	202	1392m	0.054	ng/ul	
17) Pyrene	20.06	202	1187m	0.041	ng/ul	
18) Benzo(a)anthracene	21.77	228	976	0.040	ng/ul#	74
19) Chrysene	21.83	228	667	0.027	ng/ul#	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
Data File : BE095873.D
Acq On : 28 Mar 2018 12:59
Operator : SJ/JU
Sample : J2060-07
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C0AT3

Manual Integrations
APPROVED

Sohil
3/29/2018 5:17:34 PM

Quant Time: Mar 29 11:28:41 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 28 16:31:37 2018
Response via : Initial Calibration

