

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
 Data File : BE095881.D
 Acq On : 28 Mar 2018 17:43
 Operator : SJ/JU
 Sample : J1950-10
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C0AK5

Manual Integrations
 APPROVED

Sohil
 3/29/2018 5:17:54 PM

Quant Time: Mar 29 11:42:15 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 29 04:43:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.41	152	1700	0.40	ng/ul	0.00
2) Naphthalene-d8	11.24	136	4494	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.00	164	3009	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.70	188	7306m	0.40	ng/ul	0.00
16) Chrysene-d12	21.79	240	7686m	0.40	ng/ul	0.00
20) Perylene-d12	24.42	264	7781m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.80	152	2749	0.36	ng/ul	0.00
14) Fluoranthene-d10	19.68	212	8985m	0.39	ng/ul	0.00
Target Compounds						
11) Pentachlorophenol	17.36	266	284m	0.277	ng/ul	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
Data File : BE095881.D
Acq On : 28 Mar 2018 17:43
Operator : SJ/JU
Sample : J1950-10
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C0AK5

Manual Integrations
APPROVED

Sohil
3/29/2018 5:17:54 PM

Quant Time: Mar 29 11:42:15 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 29 04:43:16 2018
Response via : Initial Calibration

