

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
Data File : BE095887.D
Acq On : 28 Mar 2018 21:18
Operator : SJ/JU
Sample : J2059-02
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C0AK1

Manual Integrations
APPROVED

Sohil
3/29/2018 5:18:12 PM

Quant Time: Mar 29 05:39:37 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 29 04:43:20 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.42	152	1748	0.40	ng/ul	0.00
2) Naphthalene-d8	11.24	136	4993	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.00	164	3351	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	8023m	0.40	ng/ul	0.00
16) Chrysene-d12	21.79	240	8130m	0.40	ng/ul	0.00
20) Perylene-d12	24.41	264	8585m	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.79	152	2966	0.35	ng/ul	0.00
14) Fluoranthene-d10	19.68	212	9756m	0.38	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
Data File : BE095887.D
Acq On : 28 Mar 2018 21:18
Operator : SJ/JU
Sample : J2059-02
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C0AK1

Manual Integrations
APPROVED

Sohil
3/29/2018 5:18:12 PM

Quant Time: Mar 29 05:39:37 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 29 04:43:20 2018
Response via : Initial Calibration

