

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
 Data File : BE095888.D
 Acq On : 28 Mar 2018 21:54
 Operator : SJ/JU
 Sample : J2059-03
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C0AK2

Manual Integrations
 APPROVED

Sohil
 3/29/2018 5:18:16 PM

Quant Time: Mar 29 05:51:45 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 29 04:43:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.42	152	1660	0.40	ng/ul	0.00
2) Naphthalene-d8	11.24	136	4625	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.00	164	3152	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.70	188	7425m	0.40	ng/ul	0.00
16) Chrysene-d12	21.79	240	7217m	0.40	ng/ul	0.00
20) Perylene-d12	24.41	264	8973	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.79	152	2706	0.35	ng/ul	0.00
14) Fluoranthene-d10	19.68	212	8320m	0.35	ng/ul	0.00
Target Compounds					Ovalue	
3) Naphthalene	11.29	128	84715	6.561	ng/ul#	88
5) 2-Methylnaphthalene	12.87	142	3919	0.437	ng/ul	99
7) Acenaphthylene	14.73	152	2192	0.138	ng/ul#	89
8) Acenaphthene	15.06	153	107342	8.956	ng/ul	96
9) Fluorene	16.02	166	23251	1.788	ng/ul#	90
12) Phenanthrene	17.74	178	7251m	0.326	ng/ul	
13) Anthracene	17.83	178	3426m	0.165	ng/ul	
15) Fluoranthene	19.71	202	32944m	1.223	ng/ul	
17) Pyrene	20.07	202	21905	0.799	ng/ul#	78
18) Benzo(a)anthracene	21.77	228	9574m	0.411	ng/ul	
19) Chrysene	21.83	228	8463	0.357	ng/ul	95
21) Benzo(b)fluoranthene	23.60	252	12942m	0.416	ng/ul	
22) Benzo(k)fluoranthene	23.65	252	4695m	0.159	ng/ul	
23) Benzo(a)pyrene	24.30	252	7793	0.275	ng/ul#	81
24) Indeno(1,2,3-cd)pyrene	27.21	276	4674	0.149	ng/ul#	74
25) Dibenzo(a,h)anthracene	27.22	278	1389	0.055	ng/ul#	83
26) Benzo(a,h,i)perylene	28.09	276	5000	0.185	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032818\
Data File : BE095888.D
Acq On : 28 Mar 2018 21:54
Operator : SJ/JU
Sample : J2059-03
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C0AK2

Manual Integrations
APPROVED

Sohil
3/29/2018 5:18:16 PM

Quant Time: Mar 29 05:51:45 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE031318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 29 04:43:20 2018
Response via : Initial Calibration

