

Data Path : Z:\HPCHEM1\BNA E\DATA\BE032917\
 Data File : BE092868.D
 Acq On : 29 Mar 2017 16:08
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 30 01:42:35 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 29 17:47:56 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	-0.02
2	1,4-Dioxane	0.668	0.734	-9.9	104	-0.01
3	n-Nitrosodimethylamine	0.691	0.697	-0.9	106	0.00
4 S	2-Fluorophenol	1.164	1.068	8.2	91	0.00
5 S	Phenol-d6	1.708	1.514	11.4	92	0.00
6	bis(2-Chloroethyl)ether	1.625	1.658	-2.0	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.287	0.292	-1.7	112	0.00
9	Naphthalene	1.095	1.102	-0.6	99	0.00
10	Hexachlorobutadiene	0.147	0.148	-0.7	100	0.00
11	2-Methylnaphthalene	0.695	0.672	3.3	97	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
13 S	2,4,6-Tribromophenol	0.078	0.066	15.4	93	0.00
14 S	2-Fluorobiphenyl	1.648	1.685	-2.2	101	-0.01
15	Acenaphthylene	7.666	6.711	12.5	97	0.00
16	Acenaphthene	1.367	1.306	4.5	98	0.00
17	Fluorene	1.629	1.550	4.8	96	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
19	Hexachlorobenzene	0.185	0.187	-1.1	104	0.00
20	Pentachlorophenol	0.048	0.036	25.0#	89	0.00
21	Phenanthrene	1.217	1.336	-9.8	107	0.00
22	Anthracene	1.106	1.078	2.5	106	0.00
23	Fluoranthene	5.597	4.865	13.1	95	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	101	-0.02
25	Pyrene	4.972	4.399	11.5	96	0.00
26 S	Terphenyl-d14	0.864	0.779	9.8	95	-0.02
27	Benzo(a)anthracene	1.185	1.092	7.8	102	0.00
28	Chrysene	1.250	1.184	5.3	97	0.00
29	Bis(2-ethylhexyl)phthalate	0.362	0.249	31.2#	88	-0.02
30	Indeno(1,2,3-cd)pyrene	4.702	3.978	15.4	92	-0.02
31 I	Perylene-d12	1.000	1.000	0.0	102	0.00
32	Benzo(b)fluoranthene	1.331	1.097	17.6	87	-0.01
33	Benzo(k)fluoranthene	1.192	1.093	8.3	99	-0.01
34 C	Benzo(a)pyrene	1.100	0.870	20.9#	87	0.00
35	Dibenzo(a,h)anthracene	1.170	1.007	13.9	93	-0.02
36	Benzo(g,h,i)perylene	4.851	4.230	12.8	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1