

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE032918\
 Data File : BE095917.D
 Acq On : 29 Mar 2018 16:10
 Operator : SJ/JU
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.428

Quant Time: Mar 29 18:08:07 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE031318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 29 16:32:29 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
3	Naphthalene	1.117	1.019	8.8	94	0.00
4 SURR	2-Methylnaphthalene-d10	0.677	0.729	-7.7	112	0.00
5	2-Methylnaphthalene	0.776	0.731	5.8	99	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
7	Acenaphthylene	2.020	1.927	4.6	108	0.00
8 C	Acenaphthene	1.521	1.373	9.7	103	0.00
9	Fluorene	1.651	1.500	9.1	103	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
11	Pentachlorophenol	0.056	0.052	7.1	134	0.00
12	Phenanthrene	1.198	1.070	10.7	100	0.00
13	Anthracene	1.120	1.077	3.8	108	0.00
14 SURR	Fluoranthene-d10	1.271	1.386	-9.0	123	0.00
15 C	Fluoranthene	1.451	1.333	8.1	103	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	112	0.00
17	Pyrene	1.519	1.579	-3.9	115	0.00
18	Benzo(a)anthracene	1.291	1.247	3.4	107	0.00
19	Chrysene	1.314	1.196	9.0	100	0.00
20 I	Perylene-d12	1.000	1.000	0.0	122	0.00
21	Benzo(b)fluoranthene	1.386	1.165	15.9	102	0.00
22	Benzo(k)fluoranthene	1.317	1.160	11.9	107	0.00
23 C	Benzo(a)pyrene	1.265	1.133	10.4	110	0.00
24	Indeno(1,2,3-cd)pyrene	1.403	1.316	6.2	115	0.00
25	Dibenzo(a,h)anthracene	1.122	1.064	5.2	115	0.00
26	Benzo(g,h,i)perylene	1.205	1.079	10.5	109	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0