

Data Path : Z:\HPCHEM1\BNA E\DATA\BE033017\
 Data File : BE092870.D
 Acq On : 30 Mar 2017 9:53
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 31 01:42:03 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 29 17:47:56 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	-0.02
2	1,4-Dioxane	0.668	0.661	1.0	97	0.00
3	n-Nitrosodimethylamine	0.691	0.686	0.7	107	0.00
4 S	2-Fluorophenol	1.164	1.057	9.2	92	0.00
5 S	Phenol-d6	1.708	1.459	14.6	91	0.00
6	bis(2-Chloroethyl)ether	1.625	1.635	-0.6	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.287	0.275	4.2	105	0.00
9	Naphthalene	1.095	1.110	-1.4	99	0.00
10	Hexachlorobutadiene	0.147	0.146	0.7	98	0.00
11	2-Methylnaphthalene	0.695	0.671	3.5	97	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
13 S	2,4,6-Tribromophenol	0.078	0.065	16.7	91	0.00
14 S	2-Fluorobiphenyl	1.648	1.669	-1.3	98	-0.01
15	Acenaphthylene	7.666	6.558	14.5	94	-0.02
16	Acenaphthene	1.367	1.315	3.8	98	0.00
17	Fluorene	1.629	1.545	5.2	94	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	97	-0.02
19	Hexachlorobenzene	0.185	0.195	-5.4	104	0.00
20	Pentachlorophenol	0.048	0.039	18.8	93	-0.02
21	Phenanthrene	1.217	1.408	-15.7	109	0.00
22	Anthracene	1.106	1.070	3.3	101	0.00
23	Fluoranthene	5.597	4.888	12.7	91	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	95	-0.02
25	Pyrene	4.972	4.397	11.6	90	0.00
26 S	Terphenyl-d14	0.864	0.747	13.5	85	-0.02
27	Benzo(a)anthracene	1.185	1.148	3.1	101	0.00
28	Chrysene	1.250	1.229	1.7	95	0.00
29	Bis(2-ethylhexyl)phthalate	0.362	0.233	35.6#	77	-0.02
30	Indeno(1,2,3-cd)pyrene	4.702	4.152	11.7	89	-0.02
31 I	Perylene-d12	1.000	1.000	0.0	98	-0.01
32	Benzo(b)fluoranthene	1.331	1.109	16.7	84	-0.01
33	Benzo(k)fluoranthene	1.192	1.072	10.1	93	-0.01
34 C	Benzo(a)pyrene	1.100	0.881	19.9	85	0.00
35	Dibenzo(a,h)anthracene	1.170	1.027	12.2	91	-0.02
36	Benzo(g,h,i)perylene	4.851	4.254	12.3	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0