

Data Path : Z:\HPCHEM1\BNA E\DATA\BE033017\
 Data File : BE092876.D
 Acq On : 30 Mar 2017 15:51
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 31 01:34:13 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 29 17:47:56 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	105	0.00
2	1,4-Dioxane	0.400	0.423	-5.7	108	-0.01
3	n-Nitrosodimethylamine	0.400	0.407	-1.7	115	0.00
4 S	2-Fluorophenol	0.400	0.348	13.0	93	0.00
5 S	Phenol-d6	0.400	0.360	10.0	101	0.00
6	bis(2-Chloroethyl)ether	0.400	0.414	-3.5	110	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	106	0.00
8 S	Nitrobenzene-d5	0.400	0.395	1.3	116	0.00
9	Naphthalene	0.400	0.413	-3.2	109	0.00
10	Hexachlorobutadiene	0.400	0.395	1.3	105	0.00
11	2-Methylnaphthalene	0.400	0.394	1.5	106	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	107	0.00
13 S	2,4,6-Tribromophenol	0.400	0.324	19.0	97	0.00
14 S	2-Fluorobiphenyl	0.400	0.407	-1.7	108	-0.01
15	Acenaphthylene	0.400	0.341	14.8	102	0.00
16	Acenaphthene	0.400	0.381	4.8	106	0.00
17	Fluorene	0.400	0.384	4.0	104	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	107	-0.02
19	Hexachlorobenzene	0.400	0.424	-6.0	116	0.00
20	Pentachlorophenol	0.400	0.462	-15.5	95	0.00
21	Phenanthrene	0.400	0.465	-16.3	121	0.00
22	Anthracene	0.400	0.412	-3.0	119	0.00
23	Fluoranthene	0.400	0.349	12.8	101	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	102	-0.02
25	Pyrene	0.400	0.364	9.0	99	0.00
26 S	Terphenyl-d14	0.400	0.373	6.8	99	-0.02
27	Benzo(a)anthracene	0.400	0.355	11.3	99	0.00
28	Chrysene	0.400	0.415	-3.7	107	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.306	23.5	80	-0.02
30	Indeno(1,2,3-cd)pyrene	0.400	0.328	18.0	89	-0.02
31 I	Perylene-d12	5.000	5.000	0.0	101	0.00
32	Benzo(b)fluoranthene	0.400	0.330	17.5	87	-0.01
33	Benzo(k)fluoranthene	0.400	0.368	8.0	99	-0.01
34 C	Benzo(a)pyrene	0.400	0.316	21.0#	87	0.00
35	Dibenzo(a,h)anthracene	0.400	0.340	15.0	91	-0.02
36	Benzo(g,h,i)perylene	0.400	0.337	15.8	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1