

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE033116\
 Data File : BE091573.D
 Acq On : 31 Mar 2016 2:56
 Operator : UM/SJ
 Sample : H1946-09
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MW-2

Manual Integrations
 APPROVED

Sohil
 3/31/2016 4:56:35 PM

Quant Time: Mar 31 04:34:29 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE033016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 30 18:45:59 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	37966	5.00	ng	0.00
6) Naphthalene-d8	10.59	136	187646	5.00	ng	0.00
10) Acenaphthene-d10	14.42	164	120585	5.00	ng	-0.03
16) Phenanthrene-d10	17.16	188	279814	5.00	ng	-0.01
22) Chrysene-d12	21.31	240	379166	5.00	ng	-0.02
28) Perylene-d12	23.76	264	318772	5.00	ng	-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.41	112	23615	2.38	ng	-0.01
5) Phenol-d6	6.97	99	22728	1.69	ng	-0.02
7) Nitrobenzene-d5	8.96	82	40719	3.23	ng	-0.02
11) 2,4,6-Tribromophenol	15.91	330	33431	6.26	ng	-0.01
12) 2-Fluorobiphenyl	13.05	172	112976	3.30	ng	-0.02
24) Terphenyl-d14	19.76	244	188829	3.92	ng	-0.02
Target Compounds						
8) Naphthalene	10.63	128	27632m	0.71	ng	Qvalue
9) 2-Methylnaphthalene	12.24	142	8776	0.35	ng	# 80
14) Acenaphthene	14.49	154	19004	0.64	ng	93
15) Fluorene	15.47	166	6410	0.17	ng	# 95
19) Phenanthrene	17.19	178	4603	0.07	ng	# 95
20) Anthracene	17.29	178	6487	0.11	ng	# 92
21) Fluoranthene	19.21	202	2561m	0.03	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE033116\
Data File : BE091573.D
Acq On : 31 Mar 2016 2:56
Operator : UM/SJ
Sample : H1946-09
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MW-2

Manual Integrations
APPROVED

Sohil
3/31/2016 4:56:35 PM

Quant Time: Mar 31 04:34:29 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE033016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 30 18:45:59 2016
Response via : Initial Calibration

