

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE033116\
 Data File : BE091564.D
 Acq On : 30 Mar 2016 21:22
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 31 00:55:58 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE033016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 30 18:45:59 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	110	0.00
2	1,4-Dioxane	0.400	0.370	7.5	104	-0.01
3	n-Nitrosodimethylamine	0.400	0.387	3.3	104	-0.03
4 S	2-Fluorophenol	0.400	0.376	6.0	103	-0.01
5 S	Phenol-d6	0.400	0.367	8.3	104	-0.02
6 I	Naphthalene-d8	5.000	5.000	0.0	109	0.00
7 S	Nitrobenzene-d5	0.400	0.370	7.5	103	-0.02
8	Naphthalene	0.400	0.393	1.8	104	-0.02
9	2-Methylnaphthalene	0.400	0.386	3.5	104	-0.02
10 I	Acenaphthene-d10	5.000	5.000	0.0	104	-0.03
11 S	2,4,6-Tribromophenol	0.400	0.295	26.3#	65	-0.01
12 S	2-Fluorobiphenyl	0.400	0.405	-1.3	103	-0.02
13	Acenaphthylene	0.400	0.405	-1.3	117	-0.01
14	Acenaphthene	0.400	0.395	1.3	103	-0.01
15	Fluorene	0.400	0.388	3.0	101	-0.01
16 I	Phenanthrene-d10	5.000	5.000	0.0	102	-0.01
17	Hexachlorobenzene	0.400	0.395	1.3	101	-0.02
18	Pentachlorophenol	0.400	0.379	5.3	111	-0.02
19	Phenanthrene	0.400	0.394	1.5	99	-0.01
20	Anthracene	0.400	0.373	6.8	97	-0.01
21	Fluoranthene	0.400	0.377	5.8	97	-0.02
22 I	Chrysene-d12	5.000	5.000	0.0	110	-0.02
23	Pyrene	0.400	0.353	11.8	96	0.00
24 S	Terphenyl-d14	0.400	0.357	10.8	96	-0.02
25	Benzo(a)anthracene	0.400	0.376	6.0	101	-0.02
26	Chrysene	0.400	0.366	8.5	98	-0.02
27	Indeno(1,2,3-cd)pyrene	0.400	0.353	11.8	98	-0.06
28 I	Perylene-d12	5.000	5.000	0.0	102	-0.03
29	Benzo(b)fluoranthene	0.400	0.399	0.3	101	-0.02
30	Benzo(k)fluoranthene	0.400	0.364	9.0	95	-0.02
31 C	Benzo(a)pyrene	0.400	0.374	6.5	97	-0.03
32	Dibenzo(a,h)anthracene	0.400	0.378	5.5	97	-0.05
33	Benzo(g,h,i)perylene	0.400	0.378	5.5	97	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0