

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE033116\
 Data File : BE091576.D
 Acq On : 31 Mar 2016 4:46
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 31 05:17:12 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE033016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 30 18:45:59 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	103	0.00
2	1,4-Dioxane	0.400	0.384	4.0	100	-0.02
3	n-Nitrosodimethylamine	0.400	0.392	2.0	99	-0.03
4 S	2-Fluorophenol	0.400	0.395	1.3	101	-0.02
5 S	Phenol-d6	0.400	0.390	2.5	104	-0.02
6 I	Naphthalene-d8	5.000	5.000	0.0	104	0.00
7 S	Nitrobenzene-d5	0.400	0.387	3.3	103	-0.02
8	Naphthalene	0.400	0.401	-0.3	101	-0.02
9	2-Methylnaphthalene	0.400	0.395	1.3	102	-0.01
10 I	Acenaphthene-d10	5.000	5.000	0.0	107	-0.03
11 S	2,4,6-Tribromophenol	0.400	0.314	21.5	71	-0.01
12 S	2-Fluorobiphenyl	0.400	0.388	3.0	101	-0.02
13	Acenaphthylene	0.400	0.404	-1.0	121	-0.01
14	Acenaphthene	0.400	0.397	0.8	107	-0.01
15	Fluorene	0.400	0.392	2.0	105	-0.02
16 I	Phenanthrene-d10	5.000	5.000	0.0	108	-0.01
17	Hexachlorobenzene	0.400	0.391	2.3	106	-0.02
18	Pentachlorophenol	0.400	0.366	8.5	114	-0.03
19	Phenanthrene	0.400	0.397	0.8	106	-0.01
20	Anthracene	0.400	0.385	3.8	106	-0.01
21	Fluoranthene	0.400	0.393	1.8	107	-0.02
22 I	Chrysene-d12	5.000	5.000	0.0	110	-0.02
23	Pyrene	0.400	0.394	1.5	107	0.00
24 S	Terphenyl-d14	0.400	0.382	4.5	102	-0.02
25	Benzo(a)anthracene	0.400	0.413	-3.2	111	-0.02
26	Chrysene	0.400	0.409	-2.2	110	-0.02
27	Indeno(1,2,3-cd)pyrene	0.400	0.374	6.5	104	-0.05
28 I	Perylene-d12	5.000	5.000	0.0	107	-0.03
29	Benzo(b)fluoranthene	0.400	0.395	1.3	105	-0.02
30	Benzo(k)fluoranthene	0.400	0.394	1.5	108	-0.02
31 C	Benzo(a)pyrene	0.400	0.398	0.5	108	-0.03
32	Dibenzo(a,h)anthracene	0.400	0.387	3.3	104	-0.06
33	Benzo(g,h,i)perylene	0.400	0.388	3.0	104	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0