

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE033121\
 Data File : BE101170.D
 Acq On : 31 Mar 2021 20:05
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Apr 01 01:41:10 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE032421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 31 11:11:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.850	152	4280	0.40	ng	0.00
7) Naphthalene-d8	10.628	136	19144	0.40	ng	#-0.01
13) Acenaphthene-d10	14.458	164	12632	0.40	ng	-0.01
19) Phenanthrene-d10	17.210	188	31087	0.40	ng	#-0.02
29) Chrysene-d12	21.367	240	37608	0.40	ng	#-0.01
36) Perylene-d12	23.849	264	37391	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.432	112	5254	0.52	ng	0.00
5) Phenol-d6	6.998	99	8015	0.51	ng	0.00
8) Nitrobenzene-d5	8.978	82	5658	0.50	ng	-0.01
11) 2-Methylnaphthalene-d10	12.210	152	12843	0.41	ng	0.00
14) 2,4,6-Tribromophenol	15.955	330	1553	0.69	ng	-0.02
15) 2-Fluorobiphenyl	13.089	172	17816	0.38	ng	-0.01
27) Fluoranthene-d10	19.226	212	156191	0.40	ng	0.00
31) Terphenyl-d14	19.824	244	34216	0.40	ng	-0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.382	88	3132	0.45	ng	98
3) n-Nitrosodimethylamine	3.681	42	3242	0.45	ng	92
6) bis(2-Chloroethyl)ether	7.263	93	5664	0.43	ng	95
9) Naphthalene	10.680	128	79157	0.41	ng	99
10) Hexachlorobutadiene	10.966	225	3344	0.38	ng	99
12) 2-Methylnaphthalene	12.282	142	13711	0.41	ng	98
16) Acenaphthylene	14.184	152	18649	0.38	ng	100
17) Acenaphthene	14.515	154	14036	0.40	ng	96
18) Fluorene	15.517	166	17648	0.39	ng	100
20) 4,6-Dinitro-2-methylph...	15.593	198	1129	0.72	ng	92
21) 4-Bromophenyl-phenylether	16.409	248	5779	0.41	ng	# 75
22) Hexachlorobenzene	16.530	284	5876	0.39	ng	98
23) Atrazine	16.681	200	5066	0.58	ng	# 92
24) Pentachlorophenol	16.863	266	2254	0.56	ng	99
25) Phenanthrene	17.256	178	33429	0.39	ng	99
26) Anthracene	17.346	178	28525	0.39	ng	100
28) Fluoranthene	19.255	202	42555	0.39	ng	99
30) Pyrene	19.617	202	44284	0.40	ng	99
32) Benzo(a)anthracene	21.354	228	43639	0.45	ng	96
33) Chrysene	21.403	228	47123	0.41	ng	98
34) Bis(2-ethylhexyl)phtha...	21.294	149	53165	0.84	ng	99
35) Indeno(1,2,3-cd)pyrene	26.470	276	42679	0.56	ng	99
37) Benzo(b)fluoranthene	23.086	252	42614	0.43	ng	100
38) Benzo(k)fluoranthene	23.086	252	42614	0.34	ng	99
39) Benzo(a)pyrene	23.738	252	36701	0.40	ng	98
40) Dibenzo(a,h)anthracene	26.498	278	35709	0.48	ng	99
41) Benzo(g,h,i)perylene	27.272	276	39057	0.41	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE033121\
Data File : BE101170.D
Acq On : 31 Mar 2021 20:05
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Apr 01 01:41:10 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE032421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 31 11:11:33 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.850 min Scan# 433
 Delta R.T. -0.001 min
 Lab File: BE101170.D
 Acq: 31 Mar 2021 20:05

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion: 152 Resp: 4280




Ion	Ratio	Lower	Upper
152	100		
150	153.7	126.3	189.5
115	58.2	48.9	73.3




#2
 1,4-Dioxane
 Concen: 0.45 ng
 RT: 3.382 min Scan# 45
 Delta R.T. -0.001 min
 Lab File: BE101170.D
 Acq: 31 Mar 2021 20:05

Tgt Ion: 88 Resp: 3132

Ion	Ratio	Lower	Upper
88	100		
58	88.4	68.5	102.7
43	33.7	27.2	40.8









#3
n-Nitrosodimethylamine
Concen: 0.45 ng
RT: 3.681 min Scan# 71
Delta R.T. -0.001 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	42	Resp:	3242
Ion	Ratio	Lower	Upper
42	100		
74	117.8	87.0	130.6
44	12.2	8.3	12.5



#4
2-Fluorophenol
Concen: 0.52 ng
RT: 5.432 min Scan# 223
Delta R.T. -0.001 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05

Tgt Ion:	112	Resp:	5254
Ion	Ratio	Lower	Upper
112	100		
64	77.4	60.8	91.2
63	149.2	119.4	179.0









#5
Phenol-d6
Concen: 0.51 ng
RT: 6.998 min Scan# 359
Delta R.T. -0.001 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	99	Resp:	8015
Ion	Ratio	Lower	Upper
99	100		
42	23.8	18.8	28.2
71	33.0	26.3	39.5



#6
bis(2-Chloroethyl)ether
Concen: 0.43 ng
RT: 7.263 min Scan# 382
Delta R.T. -0.012 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05

Tgt Ion:	93	Resp:	5664
Ion	Ratio	Lower	Upper
93	100		
63	376.8	292.5	438.7
95	31.9	25.8	38.6









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.628 min Scan# 651
Delta R.T. -0.012 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	19144
Ion Ratio	Lower	Upper	
136	100		
137	10.7	8.9	13.3
54	39.7	20.2	30.4#
68	7.5	4.0	6.0#



#8
Nitrobenzene-d5
Concen: 0.50 ng
RT: 8.978 min Scan# 524
Delta R.T. -0.012 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05

Tgt Ion:	82	Resp:	5658
Ion Ratio	Lower	Upper	
82	100		
128	104.1	101.5	152.3
54	267.7	189.8	284.6









#9
Naphthalene
Concen: 0.41 ng
RT: 10.680 min Scan# 655
Delta R.T. 0.001 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	128	Resp:	79157
Ion Ratio	Lower	Upper	
128	100		
129	2.9	2.2	3.2
127	3.3	2.7	4.1



#10
Hexachlorobutadiene
Concen: 0.38 ng
RT: 10.966 min Scan# 677
Delta R.T. -0.012 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05

Tgt Ion:	225	Resp:	3344
Ion Ratio	Lower	Upper	
225	100		
223	63.5	50.1	75.1
227	64.1	51.7	77.5









#11
2-Methylnaphthalene-d10
Concen: 0.41 ng
RT: 12.210 min Scan# 764
Delta R.T. 0.000 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 12843
Ion Ratio Lower Upper
152 100
151 20.2 16.6 24.8



#12
2-Methylnaphthalene
Concen: 0.41 ng
RT: 12.282 min Scan# 769
Delta R.T. 0.000 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05

Tgt Ion:142 Resp: 13711
Ion Ratio Lower Upper
142 100
141 88.5 71.8 107.6
115 33.0 27.8 41.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.458 min Scan# 920
Delta R.T. -0.014 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164	Resp:	12632
Ion Ratio	Lower	Upper
164	100	
162	103.7	76.6 115.0
160	45.0	34.5 51.7



#14
2,4,6-Tribromophenol
Concen: 0.69 ng
RT: 15.955 min Scan# 1016
Delta R.T. -0.015 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05

Tgt Ion:330	Resp:	1553
Ion Ratio	Lower	Upper
330	100	
332	97.6	69.1 103.7
141	40.4	30.6 45.8









#15
2-Fluorobiphenyl
Concen: 0.38 ng
RT: 13.089 min Scan# 825
Delta R.T. -0.014 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	172	Resp:	17816
Ion Ratio	Lower	Upper	
172	100		
171	34.6	26.4	39.6
170	23.7	17.4	26.0



#16
Acenaphthylene
Concen: 0.38 ng
RT: 14.184 min Scan# 901
Delta R.T. 0.000 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05

Tgt Ion:	152	Resp:	18649
Ion Ratio	Lower	Upper	
152	100		
151	19.2	15.4	23.0
153	13.0	10.3	15.5





#17

Acenaphthene

Concen: 0.40 ng

RT: 14.515 min Scan# 924

Delta R.T. -0.014 min

Lab File: BE101170.D

Acq: 31 Mar 2021 20:05

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:154 Resp: 14036

Ion Ratio Lower Upper

154 100

153 106.2 86.7 130.1

152 51.9 46.1 69.1



#18

Fluorene

Concen: 0.39 ng

RT: 15.517 min Scan# 987

Delta R.T. 0.000 min

Lab File: BE101170.D

Acq: 31 Mar 2021 20:05




Tgt Ion:166 Resp: 17648


Ion Ratio Lower Upper

166 100

165 99.0 78.9 118.3

167 13.6 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.210 min Scan# 1099
Delta R.T. -0.015 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 31087
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.2 4.8 7.2#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.72 ng
RT: 15.593 min Scan# 992
Delta R.T. 0.000 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05

Tgt Ion:198 Resp: 1129
Ion Ratio Lower Upper
198 100
51 79.5 68.8 103.2
105 34.3 32.7 49.1









#21
4-Bromophenyl-phenylether
Concen: 0.41 ng
RT: 16.409 min Scan# 1046
Delta R.T. -0.015 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 5779
Ion Ratio Lower Upper
248 100
250 90.8 79.0 118.6
141 74.4 24.9 37.3#



#22
Hexachlorobenzene
Concen: 0.39 ng
RT: 16.530 min Scan# 1054
Delta R.T. 0.000 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05

Tgt Ion:284 Resp: 5876
Ion Ratio Lower Upper
284 100
142 33.5 28.3 42.5
249 33.2 27.1 40.7





#23

Atrazine

Concen: 0.58 ng

RT: 16.681 min Scan# 1064

Delta R.T. 0.000 min

Lab File: BE101170.D

Acq: 31 Mar 2021 20:05





Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:200 Resp: 5066
 Ion Ratio Lower Upper
 200 100
 173 19.9 23.9 35.9#
 215 51.7 40.6 61.0



#24

Pentachlorophenol

Concen: 0.56 ng




RT: 16.863 min Scan# 1076


Delta R.T. -0.000 min

Lab File: BE101170.D

Acq: 31 Mar 2021 20:05

Tgt Ion:266 Resp: 2254
 Ion Ratio Lower Upper
 266 100
 264 63.3 51.4 77.0
 268 62.7 49.9 74.9









#25
Phenanthrene
Concen: 0.39 ng
RT: 17.256 min Scan# 1102
Delta R.T. -0.000 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 33429
Ion Ratio Lower Upper
178 100
176 18.9 14.7 22.1
179 15.6 12.4 18.6



#26
Anthracene
Concen: 0.39 ng
RT: 17.346 min Scan# 1108
Delta R.T. 0.000 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05

Tgt Ion:178 Resp: 28525
Ion Ratio Lower Upper
178 100
176 18.0 14.4 21.6
179 15.6 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.40 ng
RT: 19.226 min Scan# 1360
Delta R.T. -0.005 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 156191
Ion Ratio Lower Upper
212 100
106 3.3 2.6 3.8
104 1.8 1.4 2.2



#28
Fluoranthene
Concen: 0.39 ng
RT: 19.255 min Scan# 1365
Delta R.T. -0.005 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05

Tgt Ion:202 Resp: 42555
Ion Ratio Lower Upper
202 100
101 11.6 8.9 13.3
203 16.9 13.8 20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.367 min Scan# 1604
Delta R.T. -0.012 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 37608
Ion Ratio Lower Upper
240 100
120 7.1 4.1 6.1#
236 26.2 20.3 30.5



#30
Pyrene
Concen: 0.40 ng
RT: 19.617 min Scan# 1428
Delta R.T. -0.005 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05

Tgt Ion:202 Resp: 44284
Ion Ratio Lower Upper
202 100
200 20.6 16.2 24.2
203 17.2 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.40 ng
 RT: 19.824 min Scan# 1464
 Delta R.T. -0.011 min
 Lab File: BE101170.D
 Acq: 31 Mar 2021 20:05




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:244 Resp: 34216
 Ion Ratio Lower Upper
 244 100
 212 27.8 21.7 32.5
 122 10.8 6.6 10.0#



#32
 Benzo(a)anthracene
 Concen: 0.45 ng
 RT: 21.354 min Scan# 1603
 Delta R.T. 0.000 min
 Lab File: BE101170.D
 Acq: 31 Mar 2021 20:05

Tgt Ion:228 Resp: 43639
 Ion Ratio Lower Upper
 228 100
 226 25.4 21.8 32.8
 229 20.7 15.4 23.0









#33
Chrysene
Concen: 0.41 ng
RT: 21.403 min Scan# 1607
Delta R.T. -0.012 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	228	Resp:	47123
Ion Ratio	Lower	Upper	
228	100		
226	29.9	22.8	34.2
229	19.4	16.2	24.2



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.84 ng
RT: 21.294 min Scan# 1598
Delta R.T. -0.012 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05

Tgt Ion:	149	Resp:	53165
Ion Ratio	Lower	Upper	
149	100		
167	7.1	5.4	8.2
279	1.2	0.9	1.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.56 ng
RT: 26.470 min Scan# 2899
Delta R.T. -0.014 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:276 Resp: 42679
Ion Ratio Lower Upper
276 100
138 22.1 17.1 25.7
277 25.1 19.8 29.8



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.849 min Scan# 2168
Delta R.T. -0.014 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05

Tgt Ion:264 Resp: 37391
Ion Ratio Lower Upper
264 100
260 24.4 20.5 30.7
265 25.8 20.6 30.8









#37
Benzo(b)fluoranthene
Concen: 0.43 ng
RT: 23.086 min Scan# 1954
Delta R.T. -0.014 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 42614
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.8
125 9.4 7.5 11.3



#38
Benzo(k)fluoranthene
Concen: 0.34 ng
RT: 23.086 min Scan# 1954
Delta R.T. -0.064 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05

Tgt Ion:252 Resp: 42614
Ion Ratio Lower Upper
252 100
253 22.3 18.2 27.4
125 9.4 7.5 11.3









#39
Benzo(a)pyrene
Concen: 0.40 ng
RT: 23.738 min Scan# 2137
Delta R.T. -0.011 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252	Resp:	36701
Ion Ratio	Lower	Upper
252	100	
253	22.4	18.6 28.0
125	10.8	8.3 12.5



#40
Dibenzo(a,h)anthracene
Concen: 0.48 ng
RT: 26.498 min Scan# 2907
Delta R.T. -0.018 min
Lab File: BE101170.D
Acq: 31 Mar 2021 20:05

Tgt Ion:278	Resp:	35709
Ion Ratio	Lower	Upper
278	100	
139	15.4	12.0 18.0
279	24.4	20.2 30.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.41 ng
 RT: 27.272 min Scan# 3124
 Delta R.T. -0.018 min
 Lab File: BE101170.D
 Acq: 31 Mar 2021 20:05

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:	276	Resp:	39057
Ion Ratio	Lower	Upper	
276	100		
277	23.5	19.7	29.5
138	18.6	15.3	22.9

