

Data Path : Z:\svoasrv\HPCHEM1\BNA\_E\Data\BE033121\  
 Data File : BE101157.D  
 Acq On : 31 Mar 2021 12:02  
 Operator : CG/JU  
 Sample : M1867-01  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 GWBR1-102-115-032921

Quant Time: Mar 31 12:45:08 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_E\Methods\8270-SIM-BE032421.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Mar 31 11:11:33 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc | Units | Dev(Min) |
|-------------------------------|--------|------|----------|------|-------|----------|
| -----                         |        |      |          |      |       |          |
| Internal Standards            |        |      |          |      |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.850  | 152  | 4876     | 0.40 | ng    | 0.00     |
| 7) Naphthalene-d8             | 10.628 | 136  | 20672    | 0.40 | ng    | #-0.01   |
| 13) Acenaphthene-d10          | 14.458 | 164  | 12555    | 0.40 | ng    | -0.01    |
| 19) Phenanthrene-d10          | 17.211 | 188  | 31438    | 0.40 | ng    | #-0.01   |
| 29) Chrysene-d12              | 21.366 | 240  | 41258    | 0.40 | ng    | #-0.01   |
| 36) Perylene-d12              | 23.856 | 264  | 38068    | 0.40 | ng    | 0.00     |
| System Monitoring Compounds   |        |      |          |      |       |          |
| 4) 2-Fluorophenol             | 5.432  | 112  | 1837     | 0.16 | ng    | 0.00     |
| 5) Phenol-d6                  | 6.998  | 99   | 1563     | 0.09 | ng    | 0.00     |
| 8) Nitrobenzene-d5            | 8.991  | 82   | 4085     | 0.33 | ng    | 0.00     |
| 11) 2-Methylnaphthalene-d10   | 12.210 | 152  | 9901     | 0.29 | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol      | 15.956 | 330  | 1054     | 0.47 | ng    | -0.01    |
| 15) 2-Fluorobiphenyl          | 13.089 | 172  | 14451    | 0.31 | ng    | -0.01    |
| 27) Fluoranthene-d10          | 19.226 | 212  | 160836   | 0.41 | ng    | 0.00     |
| 31) Terphenyl-d14             | 19.829 | 244  | 42894    | 0.46 | ng    | 0.00     |
| Target Compounds              |        |      |          |      |       |          |
| 2) 1,4-Dioxane                | 3.382  | 88   | 3592     | 0.45 | ng    | # 64     |
| 9) Naphthalene                | 10.680 | 128  | 4947     | 0.02 | ng    | # 96     |
| 25) Phenanthrene              | 17.256 | 178  | 4253     | 0.05 | ng    | 98       |
| 28) Fluoranthene              | 19.255 | 202  | 4809     | 0.04 | ng    | # 91     |
| 30) Pyrene                    | 19.617 | 202  | 3573     | 0.03 | ng    | 95       |
| 34) Bis(2-ethylhexyl)phtha... | 21.306 | 149  | 15435    | 0.22 | ng    | 99       |
| 37) Benzo(b)fluoranthene      | 23.100 | 252  | 2013     | 0.02 | ng    | # 59     |
| -----                         |        |      |          |      |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_E\Data\BE033121\  
Data File : BE101157.D  
Acq On : 31 Mar 2021 12:02  
Operator : CG/JU  
Sample : M1867-01  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :  
GWBR1-102-115-032921

Quant Time: Mar 31 12:45:08 2021  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_E\Methods\8270-SIM-BE032421.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Mar 31 11:11:33 2021  
Response via : Initial Calibration

