

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE033121\
 Data File : BE101154.D
 Acq On : 31 Mar 2021 10:09
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 31 11:12:07 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE032421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 31 11:11:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	111	0.00
2	1,4-Dioxane	0.400	0.431	-7.7	105	0.00
3	n-Nitrosodimethylamine	0.400	0.419	-4.7	107	0.00
4 S	2-Fluorophenol	0.400	0.487	-21.7	119	0.00
5 S	Phenol-d6	0.400	0.446	-11.5	110	0.00
6	bis(2-Chloroethyl)ether	0.400	0.418	-4.5	103	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	109	-0.01
8 S	Nitrobenzene-d5	0.400	0.401	-0.3	112	0.00
9	Naphthalene	0.400	0.412	-3.0	102	0.00
10	Hexachlorobutadiene	0.400	0.382	4.5	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.391	2.3	99	0.00
12	2-Methylnaphthalene	0.400	0.395	1.3	101	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	107	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.483	-20.7	142	-0.01
15 S	2-Fluorobiphenyl	0.400	0.397	0.8	97	-0.01
16	Acenaphthylene	0.400	0.347	13.3	90	0.00
17	Acenaphthene	0.400	0.391	2.3	101	0.00
18	Fluorene	0.400	0.385	3.8	96	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	99	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.602	-50.5#	144	0.00
21	4-Bromophenyl-phenylether	0.400	0.416	-4.0	97	-0.01
22	Hexachlorobenzene	0.400	0.393	1.8	89	0.00
23	Atrazine	0.400	0.448	-12.0	112	0.00
24	Pentachlorophenol	0.400	0.453	-13.2	135	0.00
25	Phenanthrene	0.400	0.399	0.3	91	0.00
26	Anthracene	0.400	0.357	10.8	85	0.00
27 SURR	Fluoranthene-d10	0.400	0.388	3.0	88	0.00
28	Fluoranthene	0.400	0.380	5.0	85	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	107	0.00
30	Pyrene	0.400	0.368	8.0	87	0.00
31 S	Terphenyl-d14	0.400	0.389	2.8	92	0.00
32	Benzo(a)anthracene	0.400	0.412	-3.0	106	0.00
33	Chrysene	0.400	0.435	-8.7	100	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.493	-23.2	140	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.497	-24.2	129	0.00
36 I	Perylene-d12	0.400	0.400	0.0	122	0.00
37	Benzo(b)fluoranthene	0.400	0.448	-12.0	131	0.00
38	Benzo(k)fluoranthene	0.400	0.416	-4.0	118	0.00
39 C	Benzo(a)pyrene	0.400	0.396	1.0	119	0.00
40	Dibenzo(a,h)anthracene	0.400	0.444	-11.0	135	0.00
41	Benzo(g,h,i)perylene	0.400	0.477	-19.2	135	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_E

LabSampleId :

SSTDCCC0.4