

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE040116\
 Data File : BE091585.D
 Acq On : 31 Mar 2016 15:22
 Operator : UM/SJ
 Sample : H1946-12
 Misc : LOQ WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F.B

Quant Time: Apr 01 01:00:40 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE033016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 30 18:45:59 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	36889	5.00	ng	0.00
6) Naphthalene-d8	10.59	136	171324	5.00	ng	0.00
10) Acenaphthene-d10	14.42	164	104748	5.00	ng	-0.03
16) Phenanthrene-d10	17.16	188	268717	5.00	ng	-0.01
22) Chrysene-d12	21.31	240	373559	5.00	ng	-0.02
28) Perylene-d12	23.76	264	320264	5.00	ng	-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.41	112	30159	3.12	ng	-0.02
5) Phenol-d6	6.97	99	24791	1.90	ng	-0.02
7) Nitrobenzene-d5	8.96	82	44562	3.87	ng	-0.02
11) 2,4,6-Tribromophenol	15.91	330	31630	6.81	ng	-0.01
12) 2-Fluorobiphenyl	13.05	172	132028	4.43	ng	-0.02
24) Terphenyl-d14	19.76	244	231159	4.88	ng	-0.02

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE040116\
Data File : BE091585.D
Acq On : 31 Mar 2016 15:22
Operator : UM/SJ
Sample : H1946-12
Misc : LOQ WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F.B

Quant Time: Apr 01 01:00:40 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE033016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 30 18:45:59 2016
Response via : Initial Calibration

