

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE040116\
 Data File : BE091590.D
 Acq On : 31 Mar 2016 19:54
 Operator : UM/SJ
 Sample : SSTDC00.4
 Misc : LOQ WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDC00.4EC

Manual Integrations
 APPROVED

Sohil
 4/1/2016 9:40:40 PM

Quant Time: Apr 01 01:11:41 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE033016.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 30 18:45:59 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	35178	5.00	ng	0.00
6) Naphthalene-d8	10.59	136	171422	5.00	ng	0.00
10) Acenaphthene-d10	14.42	164	106396	5.00	ng	-0.03
16) Phenanthrene-d10	17.16	188	264569	5.00	ng	-0.01
22) Chrysene-d12	21.31	240	346263	5.00	ng	-0.02
28) Perylene-d12	23.76	264	312074	5.00	ng	-0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.41	112	3493	0.38	ng	-0.01
5) Phenol-d6	6.97	99	4838	0.39	ng	-0.02
7) Nitrobenzene-d5	8.96	82	4235	0.37	ng	-0.02
11) 2,4,6-Tribromophenol	15.91	330	1299	0.28	ng	-0.01
12) 2-Fluorobiphenyl	13.05	172	11671	0.39	ng	-0.02
24) Terphenyl-d14	19.76	244	17897	0.41	ng	-0.02
Target Compounds						Qvalue
2) 1,4-Dioxane	3.35	88	1852	0.35	ng	94
3) n-Nitrosodimethylamine	3.64	42	2223	0.35	ng	# 98
8) Naphthalene	10.63	128	13746	0.39	ng	97
9) 2-Methylnaphthalene	12.25	142	9081	0.39	ng	97
13) Acenaphthylene	14.14	152	18424	0.43	ng	# 93
14) Acenaphthene	14.49	154	10310	0.39	ng	96
15) Fluorene	15.47	166	13184	0.39	ng	99
17) Hexachlorobenzene	16.46	284	3881	0.40	ng	99
18) Pentachlorophenol	16.81	266	1713	0.31	ng	87
19) Phenanthrene	17.20	178	24062	0.39	ng	99
20) Anthracene	17.29	178	22205	0.39	ng	100
21) Fluoranthene	19.21	202	28344	0.39	ng	99
23) Pyrene	19.57	202	28875	0.41	ng	99
25) Benzo(a)anthracene	21.29	228	33881m	0.42	ng	
26) Chrysene	21.36	228	28819m	0.41	ng	
27) Indeno(1,2,3-cd)pyrene	26.33	276	33130	0.41	ng	100
29) Benzo(b)fluoranthene	23.01	252	30703	0.41	ng	100
30) Benzo(k)fluoranthene	23.06	252	28687	0.39	ng	96
31) Benzo(a)pyrene	23.65	252	28872	0.42	ng	98
32) Dibenzo(a,h)anthracene	26.35	278	27262	0.40	ng	98
33) Benzo(g,h,i)perylene	27.12	276	27777	0.40	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE040116\
Data File : BE091590.D
Acq On : 31 Mar 2016 19:54
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc : LOQ WATER
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

Sohil
4/1/2016 9:40:40 PM

Quant Time: Apr 01 01:11:41 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE033016.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 30 18:45:59 2016
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 5.00 ng
RT: 7.82 min Scan# 301
Delta R.T. 0.00 min
Lab File: BE091590.D
Acq: 31 Mar 2016 19:54

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
152	100		
150	132.1	122.2	183.2
115	44.5	45.9	68.9

Manual Integrations
APPROVED




Sohil
4/1/2016 9:40:40 PM




#2

1,4-Dioxane
Concen: 0.35 ng
RT: 3.35 min Scan# 44
Delta R.T. -0.01 min
Lab File: BE091590.D
Acq: 31 Mar 2016 19:54

Tgt Ion	Ratio	Lower	Upper
88	100		
58	72.4	61.0	91.6
43	37.4	25.4	38.2





#3


n-Nitrosodimethylamine
Concen: 0.35 ng
RT: 3.64 min Scan# 61
Delta R.T. -0.03 min
Lab File: BE091590.D
Acq: 31 Mar 2016 19:54

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion	Ratio	Lower	Upper
42	100		
74	106.4	84.3	126.5
44	11.2	6.7	10.1#

Manual Integrations
APPROVED

Sohil
4/1/2016 9:40:40 PM




Abundance Ion 42.00 (41.70 to 42.70): BE0
Ion 74.00 (73.70 to 74.70): BE0
Ion 44.00 (43.70 to 44.70): BE0





#4


2-Fluorophenol
Concen: 0.38 ng
RT: 5.41 min Scan# 163
Delta R.T. -0.01 min
Lab File: BE091590.D
Acq: 31 Mar 2016 19:54

Tgt Ion	Ratio	Lower	Upper
112	100		
64	69.0	53.3	79.9
63	37.4	29.8	44.8



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BE0
Ion 63.00 (62.70 to 63.70): BE0





#5

Phenol-d6

Concen: 0.39 ng

RT: 6.97 min Scan# 252

Delta R.T. -0.02 min

Lab File: BE091590.D

Acq: 31 Mar 2016 19:54

Instrument :

BNA_E


ClientSampleId :

SSTDCCC0.4EC




Tgt Ion:	99	Resp:	4838
Ion Ratio	Lower	Upper	
99	100		
42	27.0	20.6	30.8
71	36.6	29.0	43.4

Manual Integrations
APPROVED

Sohil
4/1/2016 9:40:40 PM



Abundance Ion 99.00 (98.70 to 99.70): BE0
Ion 42.00 (41.70 to 42.70): BE0
Ion 71.00 (70.70 to 71.70): BE0



#6

Naphthalene-d8

Concen: 5.00 ng

RT: 10.59 min Scan# 512




Delta R.T. -0.00 min


Lab File: BE091590.D

Acq: 31 Mar 2016 19:54

Tgt Ion:	136	Resp:	171422
Ion Ratio	Lower	Upper	
136	100		
137	11.7	8.8	13.2
54	4.5	5.2	7.8#
68	3.8	4.1	6.1#

Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BE0
Ion 68.00 (67.70 to 68.70): BE0





#7

Nitrobenzene-d5

Concen: 0.37 ng

RT: 8.96 min Scan# 392

Delta R.T. -0.02 min

Lab File: BE091590.D

Acq: 31 Mar 2016 19:54

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	82	Resp:	4235
Ion Ratio	Lower	Upper	
82	100		
128	33.7	23.5	35.3
54	59.9	52.4	78.6

Manual Integrations
APPROVED

Sohil
4/1/2016 9:40:40 PM



#8

Naphthalene

Concen: 0.39 ng




RT: 10.63 min Scan# 514


Delta R.T. -0.02 min

Lab File: BE091590.D

Acq: 31 Mar 2016 19:54

Tgt Ion:	128	Resp:	13746
Ion Ratio	Lower	Upper	
128	100		
129	11.5	8.2	12.2
127	13.7	11.8	17.8





#9

2-Methylnaphthalene

Concen: 0.39 ng

RT: 12.25 min Scan# 639

Delta R.T. -0.01 min

Lab File: BE091590.D

Acq: 31 Mar 2016 19:54

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	142	Resp:	9081
Ion Ratio	Lower	Upper	
142	100		
141	80.5	66.9	100.3
115	30.1	25.1	37.7

Manual Integrations
APPROVED

Sohil
4/1/2016 9:40:40 PM



#10

Acenaphthene-d10

Concen: 5.00 ng




RT: 14.42 min Scan# 820


Delta R.T. -0.03 min

Lab File: BE091590.D

Acq: 31 Mar 2016 19:54


Tgt Ion:	164	Resp:	106396
Ion Ratio	Lower	Upper	
164	100		
162	114.1	84.4	126.6
160	54.4	37.7	56.5





#11
 2,4,6-Tribromophenol
 Concen: 0.28 ng
 RT: 15.91 min Scan# 940
 Delta R.T. -0.01 min
 Lab File: BE091590.D
 Acq: 31 Mar 2016 19:54




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC



Tgt Ion	Ratio	Lower	Upper
330	100		
332	95.5	79.2	118.8
141	43.0	28.1	42.1#




Manual Integrations
 APPROVED


Sohil
 4/1/2016 9:40:40 PM



#12
 2-Fluorobiphenyl
 Concen: 0.39 ng
 RT: 13.05 min Scan# 708
 Delta R.T. -0.02 min
 Lab File: BE091590.D
 Acq: 31 Mar 2016 19:54

Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.4	29.8	44.8
170	23.9	21.4	32.2





#13

Acenaphthylene

Concen: 0.43 ng

RT: 14.14 min Scan# 799

Delta R.T. -0.01 min

Lab File: BE091590.D

Acq: 31 Mar 2016 19:54

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Manual Integrations
APPROVED

Sohil
4/1/2016 9:40:40 PM

Tgt Ion	Ratio	Lower	Upper
152	100		
151	22.9	16.3	24.5
153	17.7	11.1	16.7#



#14

Acenaphthene

Concen: 0.39 ng




RT: 14.49 min Scan# 825


Delta R.T. -0.01 min

Lab File: BE091590.D

Acq: 31 Mar 2016 19:54

Tgt Ion	Ratio	Lower	Upper
154	100		
153	113.6	87.3	130.9
152	55.9	42.9	64.3





#15

Fluorene

Concen: 0.39 ng

RT: 15.47 min Scan# 902

Delta R.T. -0.01 min

Lab File: BE091590.D

Acq: 31 Mar 2016 19:54

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Manual Integrations
APPROVED

Sohil
4/1/2016 9:40:40 PM

Tgt Ion:	166	Resp:	13184
Ion Ratio	Lower	Upper	
166	100		
165	98.5	79.2	118.8
167	14.1	10.8	16.2



#16

Phenanthrene-d10

Concen: 5.00 ng




RT: 17.16 min Scan# 1017


Delta R.T. -0.01 min

Lab File: BE091590.D

Acq: 31 Mar 2016 19:54

Tgt Ion:	188	Resp:	264569
Ion Ratio	Lower	Upper	
188	100		
94	6.7	5.4	8.2
80	6.0	5.0	7.4









#17
Hexachlorobenzene
Concen: 0.40 ng
RT: 16.46 min Scan# 974
Delta R.T. -0.02 min
Lab File: BE091590.D
Acq: 31 Mar 2016 19:54

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Resp	Lower	Upper
284	100	3881		
142	39.8	31.0	46.4	
249	32.0	25.8	38.8	




Manual Integrations
APPROVED


Sohil
4/1/2016 9:40:40 PM



#18
Pentachlorophenol
Concen: 0.31 ng
RT: 16.81 min Scan# 994
Delta R.T. -0.02 min
Lab File: BE091590.D
Acq: 31 Mar 2016 19:54

Tgt Ion	Ratio	Resp	Lower	Upper
266	100	1713		
268	71.2	48.2	72.2	
264	56.5	52.1	78.1	





#19

Phenanthrene

Concen: 0.39 ng

RT: 17.20 min Scan# 1020

Delta R.T. -0.01 min

Lab File: BE091590.D

Acq: 31 Mar 2016 19:54

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:	178	Resp:	24062
Ion Ratio	Lower	Upper	
178	100		
176	18.9	15.5	23.3
179	16.2	13.1	19.7

Manual Integrations
APPROVED

Sohil
4/1/2016 9:40:40 PM



#20

Anthracene

Concen: 0.39 ng




RT: 17.29 min Scan# 1028


Delta R.T. -0.01 min

Lab File: BE091590.D

Acq: 31 Mar 2016 19:54

Tgt Ion:	178	Resp:	22205
Ion Ratio	Lower	Upper	
178	100		
176	18.5	14.9	22.3
179	15.5	12.2	18.4





#21

Fluoranthene

Concen: 0.39 ng

RT: 19.21 min Scan# 1151

Delta R.T. -0.02 min

Lab File: BE091590.D

Acq: 31 Mar 2016 19:54

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tgt	Ion	202	Resp:	28344
Ion	Ratio	Lower	Upper	
202	100			
101	11.1	8.3	12.5	
203	17.4	13.9	20.9	

Manual Integrations
APPROVED

Sohil
4/1/2016 9:40:40 PM



#22

Chrysene-d12

Concen: 5.00 ng




RT: 21.31 min Scan# 1248


Delta R.T. -0.02 min

Lab File: BE091590.D

Acq: 31 Mar 2016 19:54

Tgt	Ion	240	Resp:	346263
Ion	Ratio	Lower	Upper	
240	100			
120	5.0	4.0	6.0	
236	29.4	23.0	34.4	









#23
 Pyrene
 Concen: 0.41 ng
 RT: 19.57 min Scan# 1170
 Delta R.T. 0.00 min
 Lab File: BE091590.D
 Acq: 31 Mar 2016 19:54

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion	Ratio	Resp Lower	Resp Upper
202	100		
200	21.2	16.4	24.6
203	18.1	14.6	21.8




Manual Integrations
 APPROVED


Sohil
 4/1/2016 9:40:40 PM



#24
 Terphenyl-d14
 Concen: 0.41 ng
 RT: 19.76 min Scan# 1180
 Delta R.T. -0.02 min
 Lab File: BE091590.D
 Acq: 31 Mar 2016 19:54

Tgt Ion	Ratio	Resp Lower	Resp Upper
244	100		
212	10.6	5.8	8.8#
122	18.3	1.8	2.8#









#25
Benzo(a)anthracene
Concen: 0.42 ng m
RT: 21.29 min Scan# 1247
Delta R.T. -0.02 min
Lab File: BE091590.D
Acq: 31 Mar 2016 19:54

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.8	24.0	36.0
229	17.8	13.8	20.6




Manual Integrations
APPROVED


Sohil
4/1/2016 9:40:40 PM



#26
Chrysene
Concen: 0.41 ng m
RT: 21.36 min Scan# 1250
Delta R.T. -0.00 min
Lab File: BE091590.D
Acq: 31 Mar 2016 19:54

Tgt Ion	Ratio	Lower	Upper
228	100		
226	23.4	18.9	28.3
229	26.2	19.1	28.7






#27
 Indeno(1,2,3-cd)pyrene
 Concen: 0.41 ng
 RT: 26.33 min Scan# 1647
 Delta R.T. -0.05 min
 Lab File: BE091590.D
 Acq: 31 Mar 2016 19:54

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC




Tgt Ion	Ratio	Lower	Upper
276	100		
138	25.3	20.2	30.4
277	24.4	19.8	29.8

Manual Integrations
 APPROVED

Sohil
 4/1/2016 9:40:40 PM




Abundance Ion 276.00 (275.70 to 276.70): E
 12000
 Ion 138.00 (137.70 to 138.70): E
 10000
 Ion 277.00 (276.70 to 277.70): E






#28
 Perylene-d12
 Concen: 5.00 ng
 RT: 23.76 min Scan# 1427
 Delta R.T. -0.03 min
 Lab File: BE091590.D
 Acq: 31 Mar 2016 19:54

Tgt Ion	Ratio	Lower	Upper
264	100		
260	26.2	20.1	30.1
265	21.4	18.3	27.5



Abundance Ion 264.00 (263.70 to 264.70): E
 150000
 Ion 260.00 (259.70 to 260.70): E
 100000
 Ion 265.00 (264.70 to 265.70): E









#29
Benzo(b)fluoranthene
Concen: 0.41 ng
RT: 23.01 min Scan# 1362
Delta R.T. -0.02 min
Lab File: BE091590.D
Acq: 31 Mar 2016 19:54

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.7	18.7	28.1
125	12.2	9.8	14.6




Manual Integrations
APPROVED

Sohil
4/1/2016 9:40:40 PM



#30
Benzo(k)fluoranthene
Concen: 0.39 ng
RT: 23.06 min Scan# 1366
Delta R.T. -0.02 min
Lab File: BE091590.D
Acq: 31 Mar 2016 19:54

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.3	20.2	30.4
125	11.9	9.4	14.2





#31

Benzo(a)pyrene

Concen: 0.42 ng

RT: 23.65 min Scan# 1417

Delta R.T. -0.03 min

Lab File: BE091590.D


Acq: 31 Mar 2016 19:54

Instrument :

BNA_E

ClientSampleId :




SSTDCCC0.4EC



Tgt Ion:	252	Resp:	28872
Ion Ratio	Lower	Upper	
252	100		
253	22.8	19.4	29.0
125	14.8	11.4	17.2

Manual Integrations
APPROVED

Sohil
4/1/2016 9:40:40 PM



#32

Dibenzo(a,h)anthracene

Concen: 0.40 ng




RT: 26.35 min Scan# 1649


Delta R.T. -0.05 min

Lab File: BE091590.D

Acq: 31 Mar 2016 19:54

Tgt Ion:	278	Resp:	27262
Ion Ratio	Lower	Upper	
278	100		
139	16.4	14.2	21.4
279	24.9	19.6	29.4





#33

Benzo(g,h,i)perylene

Concen: 0.40 ng

RT: 27.12 min Scan# 1715

Delta R.T. -0.05 min

Lab File: BE091590.D


Acq: 31 Mar 2016 19:54

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC



Tgt Ion:	276	Resp:	27777
Ion Ratio	Lower	Upper	
276	100		
277	24.2	19.4	29.2
138	23.6	18.2	27.4

Manual Integrations
APPROVED

Sohil
4/1/2016 9:40:40 PM

