

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040121\
 Data File : BE101172.D
 Acq On : 1 Apr 2021 11:48
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Apr 01 12:51:58 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE032421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 01 12:49:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.850	152	4258	0.40	ng	0.00
7) Naphthalene-d8	10.628	136	17904	0.40	ng	# 0.00
13) Acenaphthene-d10	14.457	164	10840	0.40	ng	0.00
19) Phenanthrene-d10	17.212	188	25487	0.40	ng	# 0.00
29) Chrysene-d12	21.366	240	31642	0.40	ng	# 0.00
36) Perylene-d12	23.852	264	29025	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.431	112	5002	0.50	ng	0.00
5) Phenol-d6	6.997	99	7654	0.49	ng	0.00
8) Nitrobenzene-d5	8.978	82	5096	0.48	ng	0.00
11) 2-Methylnaphthalene-d10	12.210	152	11672	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.957	330	1291	0.67	ng	0.00
15) 2-Fluorobiphenyl	13.088	172	16138	0.40	ng	0.00
27) Fluoranthene-d10	19.226	212	136750	0.42	ng	0.00
31) Terphenyl-d14	19.823	244	29962	0.42	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.381	88	3119	0.45	ng	98
3) n-Nitrosodimethylamine	3.681	42	3097	0.43	ng	# 97
6) bis(2-Chloroethyl)ether	7.262	93	5534	0.42	ng	99
9) Naphthalene	10.680	128	75465	0.41	ng	100
10) Hexachlorobutadiene	10.965	225	3318	0.40	ng	99
12) 2-Methylnaphthalene	12.282	142	12568	0.40	ng	97
16) Acenaphthylene	14.183	152	15742	0.38	ng	100
17) Acenaphthene	14.515	154	12092	0.40	ng	98
18) Fluorene	15.519	166	15656	0.40	ng	99
20) 4,6-Dinitro-2-methylph...	15.579	198	879	0.68	ng	# 73
21) 4-Bromophenyl-phenylether	16.411	248	5029	0.44	ng	# 84
22) Hexachlorobenzene	16.532	284	5083	0.41	ng	99
23) Atrazine	16.683	200	3907	0.55	ng	# 90
24) Pentachlorophenol	16.864	266	1658	0.50	ng	97
25) Phenanthrene	17.257	178	29229	0.41	ng	99
26) Anthracene	17.348	178	23074	0.38	ng	99
28) Fluoranthene	19.255	202	37473	0.42	ng	99
30) Pyrene	19.616	202	38010	0.41	ng	99
32) Benzo(a)anthracene	21.354	228	36514	0.45	ng	98
33) Chrysene	21.402	228	41501	0.43	ng	98
34) Bis(2-ethylhexyl)phtha...	21.293	149	30905	0.58	ng	99
35) Indeno(1,2,3-cd)pyrene	26.471	276	35155	0.55	ng	# 87
37) Benzo(b)fluoranthene	23.089	252	36292	0.48	ng	100
38) Benzo(k)fluoranthene	23.139	252	38745	0.40	ng	99
39) Benzo(a)pyrene	23.738	252	31827	0.45	ng	99
40) Dibenzo(a,h)anthracene	26.499	278	29548	0.50	ng	99
41) Benzo(g,h,i)perylene	27.270	276	33828	0.46	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040121\
 Data File : BE101172.D
 Acq On : 1 Apr 2021 11:48
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Apr 01 12:51:58 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE032421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 01 12:49:55 2021
 Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.850 min Scan# 433
 Delta R.T. -0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:152 Resp: 4258
 Ion Ratio Lower Upper
 152 100
 150 151.7 126.3 189.5
 115 58.3 48.9 73.3



1,4-Dioxane
 Concen: 0.45 ng
 RT: 3.381 min Scan# 45
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48

Tgt Ion: 88 Resp: 3119
 Ion Ratio Lower Upper
 88 100
 58 87.8 68.5 102.7
 43 32.8 27.2 40.8









n-Nitrosodimethylamine
 Concen: 0.43 ng
 RT: 3.681 min Scan# 71
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion: 42 Resp: 3097
 Ion Ratio Lower Upper
 42 100
 74 111.8 87.0 130.6
 44 6.7 8.3 12.5#



2-Fluorophenol
 Concen: 0.50 ng
 RT: 5.431 min Scan# 223
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48

Tgt Ion: 112 Resp: 5002
 Ion Ratio Lower Upper
 112 100
 64 77.8 60.8 91.2
 63 150.6 119.4 179.0









Phenol-d6
 Concen: 0.49 ng
 RT: 6.997 min Scan# 359
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	99	Resp:	7654
Ion	Ratio	Lower	Upper
99	100		
42	24.1	18.8	28.2
71	33.2	26.3	39.5




bis(2-Chloroethyl)ether
 Concen: 0.42 ng
 RT: 7.262 min Scan# 382
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48

Tgt Ion:	93	Resp:	5534
Ion	Ratio	Lower	Upper
93	100		
63	367.1	292.5	438.7
95	32.2	25.8	38.6





Abundance Scan 652 (10.640 min): BE101146.D\data.ms (-6447-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.628 min Scan# 651
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	17904
Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.9	13.3
54	36.6	20.2	30.4#
68	7.7	4.0	6.0#



Abundance Scan 651 (10.628 min): BE101172.D\data.ms (-636) (-)





Abundance Scan 525 (8.991 min): BE101146.D\data.ms (-52278)




Nitrobenzene-d5
Concen: 0.48 ng
RT: 8.978 min Scan# 524
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48


Tgt Ion:	82	Resp:	5096
Ion	Ratio	Lower	Upper
82	100		
128	99.9	101.5	152.3#
54	260.9	189.8	284.6



Abundance Scan 524 (8.978 min): BE101172.D\data.ms (-509) (-)





Abundance Scan 655 (10.679 min): BE101146.D\data.ms (-650) (-)




Naphthalene
Concen: 0.41 ng
RT: 10.680 min Scan# 655
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	128	Resp:	75465
Ion	Ratio	Lower	Upper
128	100		
129	2.9	2.2	3.2
127	3.3	2.7	4.1



Abundance Scan 655 (10.680 min): BE101172.D\data.ms (-647) (-)





Abundance Scan 678 (10.978 min): BE101146.D\data.ms (-673) (-)





Hexachlorobutadiene
Concen: 0.40 ng
RT: 10.965 min Scan# 677
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48

Tgt Ion:	225	Resp:	3318
Ion	Ratio	Lower	Upper
225	100		
223	63.0	50.1	75.1
227	64.2	51.7	77.5



Abundance Scan 677 (10.965 min): BE101172.D\data.ms (-669) (-)









2-Methylnaphthalene-d10
Concen: 0.40 ng
RT: 12.210 min Scan# 764
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 11672
Ion Ratio Lower Upper
152 100
151 20.7 16.6 24.8



2-Methylnaphthalene
Concen: 0.40 ng
RT: 12.282 min Scan# 769
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48

Tgt Ion:142 Resp: 12568
Ion Ratio Lower Upper
142 100
141 87.8 71.8 107.6
115 31.6 27.8 41.6









Acenaphthene-d10
 Concen: 0.40 ng
 RT: 14.457 min Scan# 920
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:	164	Resp:	10840
Ion	Ratio	Lower	Upper
164	100		
162	104.0	76.6	115.0
160	46.0	34.5	51.7



2,4,6-Tribromophenol
 Concen: 0.67 ng
 RT: 15.957 min Scan# 1016
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48

Tgt Ion:	330	Resp:	1291
Ion	Ratio	Lower	Upper
330	100		
332	95.4	69.1	103.7
141	37.4	30.6	45.8









2-Fluorobiphenyl
Concen: 0.40 ng
RT: 13.088 min Scan# 825
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	172	Resp:	16138
Ion	Ratio	Lower	Upper
172	100		
171	34.7	26.4	39.6
170	23.5	17.4	26.0



Acenaphthylene
Concen: 0.38 ng
RT: 14.183 min Scan# 901
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48

Tgt Ion:	152	Resp:	15742
Ion	Ratio	Lower	Upper
152	100		
151	19.4	15.4	23.0
153	12.9	10.3	15.5









Acenaphthene
 Concen: 0.40 ng
 RT: 14.515 min Scan# 924
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:	154	Resp:	12092
Ion	Ratio	Lower	Upper
154	100		
153	108.6	86.7	130.1
152	54.3	46.1	69.1



Fluorene
 Concen: 0.40 ng
 RT: 15.519 min Scan# 987
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48

Tgt Ion:	166	Resp:	15656
Ion	Ratio	Lower	Upper
166	100		
165	97.1	78.9	118.3
167	13.6	10.7	16.1









Phenanthrene-d10
 Concen: 0.40 ng
 RT: 17.212 min Scan# 1099
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:	188	Resp:	25487
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	9.8	4.8	7.2#



4,6-Dinitro-2-methylphenol
 Concen: 0.68 ng
 RT: 15.579 min Scan# 991
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48

Tgt Ion:	198	Resp:	879
Ion Ratio	Lower	Upper	
198	100		
51	106.1	68.8	103.2#
105	64.3	32.7	49.1#









Scan 1047 (16.424 min): BE101146.D\data.ms (-1024) (-)
 4-Bromophenyl-phenylether
 Concen: 0.44 ng
 RT: 16.411 min Scan# 1046
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:	248	Resp:	5029
Ion Ratio	Lower	Upper	
248	100		
250	93.2	79.0	118.6
141	57.4	24.9	37.3#



Scan 1054 (16.530 min): BE101146.D\data.ms (-1032) (-)
 Hexachlorobenzene
 Concen: 0.41 ng
 RT: 16.532 min Scan# 1054
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48

Tgt Ion:	284	Resp:	5083
Ion Ratio	Lower	Upper	
284	100		
142	34.0	28.3	42.5
249	33.9	27.1	40.7









Scan 1064 (16.681 min): BE101146.D\data.ms (-1039) (-)
 Atrazine
 Concen: 0.55 ng
 RT: 16.683 min Scan# 1064
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:	200	Resp:	3907
Ion Ratio	Lower	Upper	
200	100		
173	19.2	23.9	35.9#
215	53.3	40.6	61.0



Scan 1076 (16.863 min): BE101146.D\data.ms (-1073) (-)
 Pentachlorophenol
 Concen: 0.50 ng
 RT: 16.864 min Scan# 1076
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48

Tgt Ion:	266	Resp:	1658
Ion Ratio	Lower	Upper	
266	100		
264	61.4	51.4	77.0
268	64.4	49.9	74.9









Scan 1102 (17.256 min): BE101146.D\data.ms (-1025) (-)
 Phenanthrene
 Concen: 0.41 ng
 RT: 17.257 min Scan# 1102
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	178	Resp:	29229
Ion Ratio	Lower	Upper	
178	100		
176	18.8	14.7	22.1
179	15.5	12.4	18.6




Scan 1108 (17.346 min): BE101146.D\data.ms (-1108) (-)
 Anthracene
 Concen: 0.38 ng
 RT: 17.348 min Scan# 1108
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48

Tgt Ion:	178	Resp:	23074
Ion Ratio	Lower	Upper	
178	100		
176	18.1	14.4	21.6
179	15.0	12.2	18.4






Abundance Scan 1361 (19.232 min): BE101146.D\data.ms (-1323) (-)




Fluoranthene-d10
Concen: 0.42 ng
RT: 19.226 min Scan# 1360
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Tgt Ion:	212	Resp:	136750
Ion Ratio	Lower	Upper	
212	100		
106	3.3	2.6	3.8
104	1.8	1.4	2.2







Abundance Scan 1366 (19.261 min): BE101146.D\data.ms (-1329) (-)



Fluoranthene
Concen: 0.42 ng
RT: 19.255 min Scan# 1365
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48

Tgt Ion:	202	Resp:	37473
Ion Ratio	Lower	Upper	
202	100		
101	11.8	8.9	13.3
203	17.1	13.8	20.6









Chrysene-d12
 Concen: 0.40 ng
 RT: 21.366 min Scan# 1604
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	240	Resp:	31642
Ion Ratio	Lower	Upper	
240	100		
120	7.0	4.1	6.1#
236	26.4	20.3	30.5



Pyrene
 Concen: 0.41 ng
 RT: 19.616 min Scan# 1428
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48


Tgt Ion:	202	Resp:	38010
Ion Ratio	Lower	Upper	
202	100		
200	20.8	16.2	24.2
203	17.6	13.9	20.9






Abundance Scan 1466 (19.835 min): BE101146.D\data.ms (-1438) (-)

Terphenyl-d14
Concen: 0.42 ng
RT: 19.823 min Scan# 1464
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4




Tgt Ion	Ratio	Lower	Upper
244	100		
212	28.2	21.7	32.5
122	10.2	6.6	10.0#







Abundance Scan 1603 (21.354 min): BE101146.D\data.ms (-1584) (-)

Benzo(a)anthracene
Concen: 0.45 ng
RT: 21.354 min Scan# 1603
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48




Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.0	21.8	32.8
229	19.9	15.4	23.0








Scan 1608 (21.415 min): BE101146.D\data.ms (-1608) (-)
 Chrysene
 Concen: 0.43 ng
 RT: 21.402 min Scan# 1607
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4






Tgt Ion:	228	Resp:	41501
Ion	Ratio	Lower	Upper
228	100		
226	30.1	22.8	34.2
229	19.5	16.2	24.2



Scan 1599 (21.306 min): BE101146.D\data.ms (-1599) (-)
 Bis(2-ethylhexyl)phthalate
 Concen: 0.58 ng
 RT: 21.293 min Scan# 1598
 Delta R.T. 0.000 min
 Lab File: BE101172.D
 Acq: 1 Apr 2021 11:48


Tgt Ion:	149	Resp:	30905
Ion	Ratio	Lower	Upper
149	100		
167	7.1	5.4	8.2
279	1.2	0.9	1.3





Abundance Scan 2903 (26.484 min): BE101146.D\data.ms (-2885) (-)

Indeno(1,2,3-cd)pyrene
Concen: 0.55 ng
RT: 26.471 min Scan# 2899
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48


Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4



Tgt Ion	Ratio	Lower	Upper
276	100		
138	8.8	17.1	25.7#
277	24.9	19.8	29.8




Abundance Scan 2899 (26.471 min): BE101172.D\data.ms (-2871) (-)





Abundance Scan 2172 (23.863 min): BE101146.D\data.ms (-2130) (-)


Perylene-d12
Concen: 0.40 ng
RT: 23.852 min Scan# 2169
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48




Tgt Ion	Ratio	Lower	Upper
264	100		
260	23.8	20.5	30.7
265	26.4	20.6	30.8



Abundance Scan 2169 (23.852 min): BE101172.D\data.ms (-2127) (-)





Abundance Scan 1958 (23.100 min): BE101146.D\data.ms (-1938) (-)




Benzo(b)fluoranthene
Concen: 0.48 ng
RT: 23.089 min Scan# 1955
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48

Tgt Ion:	252	Resp:	36292
Ion Ratio	Lower	Upper	
252	100		
253	22.5	17.8	26.8
125	9.3	7.5	11.3


Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4



Abundance Scan 1955 (23.089 min): BE101172.D\data.ms (-1899) (-)





Abundance Scan 1972 (23.150 min): BE101146.D\data.ms (-1938) (-)




Benzo(k)fluoranthene
Concen: 0.40 ng
RT: 23.139 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48

Tgt Ion:	252	Resp:	38745
Ion Ratio	Lower	Upper	
252	100		
253	22.4	18.2	27.4
125	10.0	7.5	11.3




Abundance Scan 1969 (23.139 min): BE101172.D\data.ms (-1885) (-)





Abundance Scan 2140 (23.749 min): BE101146.D\data.ms (-289) (-)

Benzo(a)pyrene
Concen: 0.45 ng
RT: 23.738 min Scan# 2137
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48


Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4



Tgt Ion:	252	Resp:	31827
Ion Ratio	Lower	Upper	
252	100		
253	22.6	18.6	28.0
125	10.5	8.3	12.5





Abundance Scan 2137 (23.738 min): BE101172.D\data.ms (-2081) (-)




Abundance Scan 2912 (26.516 min): BE101146.D\data.ms (-2894) (-)

Dibenzo(a,h)anthracene
Concen: 0.50 ng
RT: 26.499 min Scan# 2907
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48


Tgt Ion:	278	Resp:	29548
Ion Ratio	Lower	Upper	
278	100		
139	14.8	12.0	18.0
279	24.7	20.2	30.4



Abundance Scan 2907 (26.499 min): BE101172.D\data.ms (-2885) (-)




Abundance Scan 3129 (27.290 min): BE101146.D\data.ms (-3104)




Benzo(g,h,i)perylene
Concen: 0.46 ng
RT: 27.270 min Scan# 3123
Delta R.T. 0.000 min
Lab File: BE101172.D
Acq: 1 Apr 2021 11:48

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Abundance Scan 3123 (27.270 min): BE101172.D\data.ms



Tgt Ion:	276	Resp:	33828
Ion	Ratio	Lower	Upper
276	100		
277	23.7	19.7	29.5
138	19.3	15.3	22.9



Abundance Scan 3123 (27.270 min): BE101172.D\data.ms (-3067)

