

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040121\
 Data File : BE101176.D
 Acq On : 1 Apr 2021 14:15
 Operator : CG/JU
 Sample : M1843-15
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 RE114D2-20210325

Quant Time: Apr 01 15:09:16 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE032421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 01 12:49:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.849	152	4114	0.40	ng	0.00
7) Naphthalene-d8	10.627	136	17850	0.40	ng	# 0.00
13) Acenaphthene-d10	14.457	164	11022	0.40	ng	0.00
19) Phenanthrene-d10	17.212	188	27986	0.40	ng	# 0.00
29) Chrysene-d12	21.366	240	32803	0.40	ng	# 0.00
36) Perylene-d12	23.848	264	29093	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.431	112	1946	0.20	ng	0.00
5) Phenol-d6	6.997	99	1706	0.11	ng	0.00
8) Nitrobenzene-d5	8.977	82	4976	0.47	ng	0.00
11) 2-Methylnaphthalene-d10	12.209	152	10943	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.957	330	747	0.38	ng	0.00
15) 2-Fluorobiphenyl	13.088	172	15971	0.39	ng	0.00
27) Fluoranthene-d10	19.225	212	172762	0.49	ng	0.00
31) Terphenyl-d14	19.823	244	41028	0.55	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.381	88	19531	2.93	ng	# 92
25) Phenanthrene	17.257	178	1913	0.02	ng	98
34) Bis(2-ethylhexyl)phtha...	21.293	149	11688	0.21	ng	# 100
40) Dibenzo(a,h)anthracene	26.498	278	612	0.11	ng	# 41

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040121\
Data File : BE101176.D
Acq On : 1 Apr 2021 14:15
Operator : CG/JU
Sample : M1843-15
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
RE114D2-20210325

Quant Time: Apr 01 15:09:16 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE032421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 01 12:49:55 2021
Response via : Initial Calibration

