

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040121\
 Data File : BE101180.D
 Acq On : 1 Apr 2021 16:44
 Operator : CG/JU
 Sample : M1843-19
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 FB01-20210325

Quant Time: Apr 01 17:28:27 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE032421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 01 12:49:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.838	152	3798	0.40	ng	-0.01
7) Naphthalene-d8	10.628	136	16064	0.40	ng	# 0.00
13) Acenaphthene-d10	14.457	164	10770	0.40	ng	0.00
19) Phenanthrene-d10	17.212	188	26845	0.40	ng	# 0.00
29) Chrysene-d12	21.366	240	34712	0.40	ng	# 0.00
36) Perylene-d12	23.845	264	30613	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.431	112	1639	0.18	ng	0.00
5) Phenol-d6	6.997	99	1365	0.10	ng	0.00
8) Nitrobenzene-d5	8.978	82	3791	0.40	ng	0.00
11) 2-Methylnaphthalene-d10	12.210	152	8480	0.32	ng	0.00
14) 2,4,6-Tribromophenol	15.957	330	228	0.12	ng	0.00
15) 2-Fluorobiphenyl	13.088	172	12078	0.30	ng	0.00
27) Fluoranthene-d10	19.226	212	149509	0.44	ng	0.00
31) Terphenyl-d14	19.823	244	34649	0.44	ng	0.00
Target Compounds						Qvalue
9) Naphthalene	10.680	128	14903	0.09	ng	# 99
12) 2-Methylnaphthalene	12.282	142	3635	0.13	ng	99
18) Fluorene	15.519	166	1199	0.03	ng	# 68
25) Phenanthrene	17.258	178	2848	0.04	ng	# 92
30) Pyrene	19.617	202	2156	0.02	ng	# 88
34) Bis(2-ethylhexyl)phtha...	21.293	149	8188	0.14	ng	98
40) Dibenzo(a,h)anthracene	26.484	278	401	0.11	ng	# 18

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040121\
Data File : BE101180.D
Acq On : 1 Apr 2021 16:44
Operator : CG/JU
Sample : M1843-19
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
FB01-20210325

Quant Time: Apr 01 17:28:27 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE032421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 01 12:49:55 2021
Response via : Initial Calibration

