

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE040218\
 Data File : BE095933.D
 Acq On : 2 Apr 2018 18:29
 Operator : SJ/JU
 Sample : J2136-02
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PK-10S-20180328

Quant Time: Apr 03 04:46:02 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 12:41:11 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.42	152	1215	0.40	ng	0.00
7) Naphthalene-d8	11.26	136	4902	0.40	ng	0.00
13) Acenaphthene-d10	14.99	164	2869	0.40	ng	-0.01
19) Phenanthrene-d10	17.69	188	6648	0.40	ng	-0.01
27) Chrysene-d12	21.79	240	7004	0.40	ng	-0.01
34) Perylene-d12	24.42	264	6961	0.40	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.91	112	612	0.15	ng	0.00
5) Phenol-d6	7.56	99	528	0.10	ng	0.00
8) Nitrobenzene-d5	9.59	82	2522	0.44	ng	0.00
11) 2-Methylnaphthalene-d10	12.79	152	3625	0.44	ng	0.00
14) 2,4,6-Tribromophenol	16.46	330	673	0.41	ng	-0.01
15) 2-Fluorobiphenyl	13.63	172	5728	0.47	ng	-0.01
25) Fluoranthene-d10	19.68	212	40149	0.41	ng	0.00
29) Terphenyl-d14	20.24	244	8453	0.57	ng	0.00
Target Compounds						Qvalue
32) Bis(2-ethylhexyl)phthalate	21.64	149	10406	0.140	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE040218\
Data File : BE095933.D
Acq On : 2 Apr 2018 18:29
Operator : SJ/JU
Sample : J2136-02
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PK-10S-20180328

Quant Time: Apr 03 04:46:02 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 12:41:11 2018
Response via : Initial Calibration

