

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE040218\
 Data File : BE095919.D
 Acq On : 2 Apr 2018 10:12
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 03 03:47:40 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 12:41:11 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.655	0.591	9.8	105	0.00
3	n-Nitrosodimethylamine	0.961	0.814	15.3	91	0.00
4 S	2-Fluorophenol	1.319	1.242	5.8	100	0.00
5 S	Phenol-d6	1.813	1.860	-2.6	109	0.00
6	bis(2-Chloroethyl)ether	1.564	1.567	-0.2	107	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	115	0.00
8 S	Nitrobenzene-d5	0.473	0.479	-1.3	115	0.00
9	Naphthalene	4.569	4.742	-3.8	115	0.00
10	Hexachlorobutadiene	0.314	0.313	0.3	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.668	0.703	-5.2	117	0.00
12	2-Methylnaphthalene	0.780	0.822	-5.4	118	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	115	-0.01
14 S	2,4,6-Tribromophenol	0.231	0.196	15.2	101	-0.01
15 S	2-Fluorobiphenyl	1.686	1.705	-1.1	114	-0.01
16	Acenaphthylene	2.288	2.367	-3.5	119	0.00
17	Acenaphthene	1.395	1.408	-0.9	117	0.00
18	Fluorene	1.835	1.813	1.2	112	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	107	-0.01
20	4-Bromophenyl-phenylether	0.240	0.245	-2.1	107	0.00
21	Hexachlorobenzene	0.240	0.246	-2.5	108	0.00
22	Pentachlorophenol	0.069	0.060	13.0	112	-0.01
23	Phenanthrene	1.197	1.222	-2.1	106	0.00
24	Anthracene	1.197	1.219	-1.8	108	0.00
25 SURR	Fluoranthene-d10	5.940	5.477	7.8	97	0.00
26	Fluoranthene	1.663	1.559	6.3	99	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	93	-0.01
28	Pvrene	1.629	1.759	-8.0	91	-0.01
29 S	Terphenyl-d14	0.850	0.865	-1.8	93	0.00
30	Benzo(a)anthracene	1.387	1.417	-2.2	92	0.00
31	Chrysene	1.266	1.332	-5.2	96	0.00
32	Bis(2-ethylhexyl)phthalate	4.242	3.968	6.5	92	0.00
33	Indeno(1,2,3-cd)pyrene	1.364	1.451	-6.4	97	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	95	-0.01
35	Benzo(b)fluoranthene	1.323	1.361	-2.9	96	0.00
36	Benzo(k)fluoranthene	1.324	1.388	-4.8	98	-0.01
37 C	Benzo(a)pyrene	1.240	1.282	-3.4	96	-0.01
38	Dibenzo(a,h)anthracene	1.178	1.207	-2.5	94	-0.02
39	Benzo(a,h,i)perylene	1.184	1.238	-4.6	97	-0.01

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA E\Data\BE040218\
Data File : BE095919.D
Acq On : 2 Apr 2018 10:12
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 03 03:47:40 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 12:41:11 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			