

Data Path : Z:\HPCHEM1\BNA E\DATA\BE040417\
 Data File : BE092886.D
 Acq On : 4 Apr 2017 15:58
 Operator : SJ/MA
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE040417

Quant Time: Apr 04 16:37:50 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 04 16:34:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	-0.02
2	1,4-Dioxane	0.671	0.684	-1.9	109	0.00
3	n-Nitrosodimethylamine	0.673	0.663	1.5	112	0.00
4 S	2-Fluorophenol	1.007	1.007	0.0	117	0.00
5 S	Phenol-d6	1.445	1.396	3.4	117	0.00
6	bis(2-Chloroethyl)ether	1.622	1.584	2.3	107	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
8 S	Nitrobenzene-d5	0.287	0.298	-3.8	126	-0.02
9	Naphthalene	1.103	1.106	-0.3	109	0.00
10	Hexachlorobutadiene	0.149	0.148	0.7	108	0.00
11	2-Methylnaphthalene	0.684	0.672	1.8	110	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	113	-0.02
13 S	2,4,6-Tribromophenol	0.064	0.059	7.8	123	-0.01
14 S	2-Fluorobiphenyl	1.689	1.663	1.5	109	-0.01
15	Acenaphthylene	7.071	6.531	7.6	117	-0.02
16	Acenaphthene	1.452	1.398	3.7	112	0.00
17	Fluorene	1.606	1.517	5.5	112	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	113	-0.02
19	Hexachlorobenzene	0.190	0.182	4.2	109	0.00
20	Pentachlorophenol	0.036	0.034	5.6	116	-0.02
21	Phenanthrene	1.335	1.297	2.8	113	0.00
22	Anthracene	1.090	0.973	10.7	114	0.00
23	Fluoranthene	4.886	4.496	8.0	111	-0.02
24 I	Chrysene-d12	1.000	1.000	0.0	110	-0.02
25	Pyrene	4.464	4.245	4.9	113	0.00
26 S	Terphenyl-d14	0.768	0.739	3.8	111	-0.02
27	Benzo(a)anthracene	1.079	0.951	11.9	114	0.00
28	Chrysene	1.252	1.283	-2.5	108	0.00
29	Bis(2-ethylhexyl)phthalate	0.260	0.223	14.2	121	-0.02
30	Indeno(1,2,3-cd)pyrene	4.172	3.710	11.1	101	-0.02
31 I	Perylene-d12	1.000	1.000	0.0	105	-0.01
32	Benzo(b)fluoranthene	1.236	1.140	7.8	104	-0.01
33	Benzo(k)fluoranthene	1.119	1.014	9.4	107	-0.01
34 C	Benzo(a)pyrene	0.961	0.831	13.5	104	-0.01
35	Dibenzo(a,h)anthracene	1.102	0.980	11.1	102	-0.03
36	Benzo(g,h,i)perylene	4.504	4.128	8.3	102	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0