

Data Path : Z:\HPCHEM1\BNA E\DATA\BE040417\
 Data File : BE092886.D
 Acq On : 4 Apr 2017 15:58
 Operator : SJ/MA
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE040417

Quant Time: Apr 04 16:37:50 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 04 16:34:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	107	-0.02
2	1,4-Dioxane	0.400	0.408	-2.0	109	0.00
3	n-Nitrosodimethylamine	0.400	0.394	1.5	112	0.00
4 S	2-Fluorophenol	0.400	0.400	0.0	117	0.00
5 S	Phenol-d6	0.400	0.386	3.5	117	0.00
6	bis(2-Chloroethyl)ether	0.400	0.391	2.3	107	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	109	0.00
8 S	Nitrobenzene-d5	0.400	0.416	-4.0	126	-0.02
9	Naphthalene	0.400	0.401	-0.3	109	0.00
10	Hexachlorobutadiene	0.400	0.397	0.8	108	0.00
11	2-Methylnaphthalene	0.400	0.393	1.8	110	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	113	-0.02
13 S	2,4,6-Tribromophenol	0.400	0.369	7.8	123	-0.01
14 S	2-Fluorobiphenyl	0.400	0.394	1.5	109	-0.01
15	Acenaphthylene	0.400	0.369	7.8	117	-0.02
16	Acenaphthene	0.400	0.385	3.8	112	0.00
17	Fluorene	0.400	0.378	5.5	112	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	113	-0.02
19	Hexachlorobenzene	0.400	0.383	4.3	109	0.00
20	Pentachlorophenol	0.400	0.430	-7.5	116	-0.02
21	Phenanthrene	0.400	0.389	2.8	113	0.00
22	Anthracene	0.400	0.357	10.8	114	0.00
23	Fluoranthene	0.400	0.368	8.0	111	-0.02
24 I	Chrysene-d12	5.000	5.000	0.0	110	-0.02
25	Pyrene	0.400	0.380	5.0	113	0.00
26 S	Terphenyl-d14	0.400	0.385	3.8	111	-0.02
27	Benzo(a)anthracene	0.400	0.352	12.0	114	0.00
28	Chrysene	0.400	0.410	-2.5	108	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.442	-10.5	121	-0.02
30	Indeno(1,2,3-cd)pyrene	0.400	0.356	11.0	101	-0.02
31 I	Perylene-d12	5.000	5.000	0.0	105	-0.01
32	Benzo(b)fluoranthene	0.400	0.369	7.8	104	-0.01
33	Benzo(k)fluoranthene	0.400	0.362	9.5	107	-0.01
34 C	Benzo(a)pyrene	0.400	0.346	13.5	104	-0.01
35	Dibenzo(a,h)anthracene	0.400	0.356	11.0	102	-0.03
36	Benzo(g,h,i)perylene	0.400	0.367	8.3	102	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0