

Data Path : Z:\HPCHEM1\BNA E\DATA\BE040417\
 Data File : BE092889.D
 Acq On : 4 Apr 2017 17:49
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 04 19:14:22 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 04 16:35:54 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	-0.02
2	1,4-Dioxane	0.671	0.667	0.6	109	0.00
3	n-Nitrosodimethylamine	0.673	0.657	2.4	114	0.00
4 S	2-Fluorophenol	1.007	0.985	2.2	117	0.00
5 S	Phenol-d6	1.445	1.403	2.9	121	0.00
6	bis(2-Chloroethyl)ether	1.622	1.620	0.1	113	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	112	-0.02
8 S	Nitrobenzene-d5	0.287	0.273	4.9	119	-0.02
9	Naphthalene	1.103	1.093	0.9	111	0.00
10	Hexachlorobutadiene	0.149	0.148	0.7	111	0.00
11	2-Methylnaphthalene	0.684	0.664	2.9	112	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	115	-0.02
13 S	2,4,6-Tribromophenol	0.064	0.059	7.8	126	-0.01
14 S	2-Fluorobiphenyl	1.689	1.678	0.7	112	-0.01
15	Acenaphthylene	7.071	6.507	8.0	119	-0.02
16	Acenaphthene	1.452	1.440	0.8	117	0.00
17	Fluorene	1.606	1.549	3.5	116	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	121	-0.02
19	Hexachlorobenzene	0.190	0.164	13.7	105	0.00
20	Pentachlorophenol	0.036	0.032	11.1	116	-0.02
21	Phenanthrene	1.335	1.217	8.8	114	0.00
22	Anthracene	1.090	0.900	17.4	113	0.00
23	Fluoranthene	4.886	4.338	11.2	115	-0.02
24 I	Chrysene-d12	1.000	1.000	0.0	112	-0.02
25	Pyrene	4.464	4.420	1.0	119	-0.02
26 S	Terphenyl-d14	0.768	0.752	2.1	114	-0.02
27	Benzo(a)anthracene	1.079	0.977	9.5	118	0.00
28	Chrysene	1.252	1.270	-1.4	108	0.00
29	Bis(2-ethylhexyl)phthalate	0.260	0.198	23.8	109	-0.02
30	Indeno(1,2,3-cd)pyrene	4.172	3.867	7.3	107	-0.02
31 I	Perylene-d12	1.000	1.000	0.0	108	-0.01
32	Benzo(b)fluoranthene	1.236	1.161	6.1	108	-0.01
33	Benzo(k)fluoranthene	1.119	1.013	9.5	110	-0.01
34 C	Benzo(a)pyrene	0.961	0.917	4.6	117	-0.01
35	Dibenzo(a,h)anthracene	1.102	0.999	9.3	107	-0.03
36	Benzo(g,h,i)perylene	4.504	4.033	10.5	102	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0