

Data Path : Z:\HPCHEM1\BNA E\DATA\BE040417\
 Data File : BE092889.D
 Acq On : 4 Apr 2017 17:49
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 04 19:14:22 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 04 16:35:54 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	110	-0.02
2	1,4-Dioxane	0.400	0.398	0.5	109	0.00
3	n-Nitrosodimethylamine	0.400	0.390	2.5	114	0.00
4 S	2-Fluorophenol	0.400	0.391	2.3	117	0.00
5 S	Phenol-d6	0.400	0.388	3.0	121	0.00
6	bis(2-Chloroethyl)ether	0.400	0.400	0.0	113	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	112	-0.02
8 S	Nitrobenzene-d5	0.400	0.381	4.8	119	-0.02
9	Naphthalene	0.400	0.396	1.0	111	0.00
10	Hexachlorobutadiene	0.400	0.397	0.8	111	0.00
11	2-Methylnaphthalene	0.400	0.389	2.8	112	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	115	-0.02
13 S	2,4,6-Tribromophenol	0.400	0.370	7.5	126	-0.01
14 S	2-Fluorobiphenyl	0.400	0.398	0.5	112	-0.01
15	Acenaphthylene	0.400	0.368	8.0	119	-0.02
16	Acenaphthene	0.400	0.397	0.8	117	0.00
17	Fluorene	0.400	0.386	3.5	116	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	121	-0.02
19	Hexachlorobenzene	0.400	0.346	13.5	105	0.00
20	Pentachlorophenol	0.400	0.410	-2.5	116	-0.02
21	Phenanthrene	0.400	0.365	8.8	114	0.00
22	Anthracene	0.400	0.330	17.5	113	0.00
23	Fluoranthene	0.400	0.355	11.3	115	-0.02
24 I	Chrysene-d12	5.000	5.000	0.0	112	-0.02
25	Pyrene	0.400	0.396	1.0	119	-0.02
26 S	Terphenyl-d14	0.400	0.392	2.0	114	-0.02
27	Benzo(a)anthracene	0.400	0.362	9.5	118	0.00
28	Chrysene	0.400	0.406	-1.5	108	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.385	3.8	109	-0.02
30	Indeno(1,2,3-cd)pyrene	0.400	0.371	7.3	107	-0.02
31 I	Perylene-d12	5.000	5.000	0.0	108	-0.01
32	Benzo(b)fluoranthene	0.400	0.376	6.0	108	-0.01
33	Benzo(k)fluoranthene	0.400	0.362	9.5	110	-0.01
34 C	Benzo(a)pyrene	0.400	0.382	4.5	117	-0.01
35	Dibenzo(a,h)anthracene	0.400	0.363	9.3	107	-0.03
36	Benzo(g,h,i)perylene	0.400	0.358	10.5	102	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0