

Data Path : Z:\HPCHEM1\BNA E\Data\BE040418\
 Data File : BE095939.D
 Acq On : 4 Apr 2018 10:06
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 05 02:46:24 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 12:41:11 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	122	-0.01
2	1,4-Dioxane	0.655	0.593	9.5	116	0.00
3	n-Nitrosodimethylamine	0.961	0.922	4.1	113	0.00
4 S	2-Fluorophenol	1.319	1.235	6.4	110	-0.01
5 S	Phenol-d6	1.813	1.917	-5.7	123	-0.01
6	bis(2-Chloroethyl)ether	1.564	1.615	-3.3	121	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	129	0.00
8 S	Nitrobenzene-d5	0.473	0.464	1.9	125	0.00
9	Naphthalene	4.569	4.793	-4.9	130	-0.01
10	Hexachlorobutadiene	0.314	0.294	6.4	115	-0.01
11 SURR	2-Methylnaphthalene-d10	0.668	0.690	-3.3	129	-0.01
12	2-Methylnaphthalene	0.780	0.826	-5.9	132	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	127	-0.01
14 S	2,4,6-Tribromophenol	0.231	0.204	11.7	116	-0.01
15 S	2-Fluorobiphenyl	1.686	1.711	-1.5	126	-0.01
16	Acenaphthylene	2.288	2.393	-4.6	133	0.00
17	Acenaphthene	1.395	1.442	-3.4	132	0.00
18	Fluorene	1.835	1.875	-2.2	128	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	120	-0.01
20	4-Bromophenyl-phenylether	0.240	0.243	-1.3	120	0.00
21	Hexachlorobenzene	0.240	0.246	-2.5	121	0.00
22	Pentachlorophenol	0.069	0.078	-13.0	162#	-0.01
23	Phenanthrene	1.197	1.251	-4.5	122	0.00
24	Anthracene	1.197	1.216	-1.6	121	-0.01
25 SURR	Fluoranthene-d10	5.940	5.559	6.4	110	0.00
26	Fluoranthene	1.663	1.589	4.4	114	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	109	-0.01
28	Pvrene	1.629	1.886	-15.8	114	-0.01
29 S	Terphenyl-d14	0.850	0.852	-0.2	108	0.00
30	Benzo(a)anthracene	1.387	1.420	-2.4	109	0.00
31	Chrysene	1.266	1.324	-4.6	112	-0.01
32	Bis(2-ethylhexyl)phthalate	4.242	3.889	8.3	106	0.00
33	Indeno(1,2,3-cd)pyrene	1.364	1.416	-3.8	111	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	110	-0.02
35	Benzo(b)fluoranthene	1.323	1.366	-3.3	112	-0.01
36	Benzo(k)fluoranthene	1.324	1.386	-4.7	114	-0.01
37 C	Benzo(a)pyrene	1.240	1.280	-3.2	111	-0.01
38	Dibenzo(a,h)anthracene	1.178	1.210	-2.7	110	-0.02
39	Benzo(a,h,i)perylene	1.184	1.234	-4.2	112	-0.02

Data Path : Z:\HPCHEM1\BNA E\Data\BE040418\
Data File : BE095939.D
Acq On : 4 Apr 2018 10:06
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 05 02:46:24 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 12:41:11 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			