

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE040419\
 Data File : BE099293.D
 Acq On : 4 Apr 2019 16:00
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 05 04:29:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 02 17:16:42 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	-0.02
2	1,4-Dioxane	0.545	0.606	-11.2	119	-0.01
3	n-Nitrosodimethylamine	0.322	0.278	13.7	95	-0.02
4 S	2-Fluorophenol	1.063	1.012	4.8	97	-0.02
5 S	Phenol-d6	1.461	1.433	1.9	103	-0.01
6	bis(2-Chloroethyl)ether	1.310	1.376	-5.0	112	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	104	-0.02
8 S	Nitrobenzene-d5	0.275	0.290	-5.5	114	-0.02
9	Naphthalene	4.351	4.679	-7.5	111	-0.02
10	Hexachlorobutadiene	0.283	0.299	-5.7	109	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.738	-5.7	112	-0.01
12	2-Methylnaphthalene	0.740	0.773	-4.5	109	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	-0.03
14 S	2,4,6-Tribromophenol	0.184	0.158	14.1	97	-0.01
15 S	2-Fluorobiphenyl	1.572	1.730	-10.1	112	-0.01
16	Acenaphthylene	1.895	1.853	2.2	103	-0.01
17	Acenaphthene	1.118	1.158	-3.6	108	-0.01
18	Fluorene	1.615	1.672	-3.5	105	-0.03
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	-0.03
20	4-Bromophenyl-phenylether	0.227	0.240	-5.7	106	-0.03
21	Hexachlorobenzene	0.216	0.239	-10.6	109	-0.03
22	Pentachlorophenol	0.090	0.092	-2.2	114	-0.03
23	Phenanthrene	1.035	1.110	-7.2	103	-0.03
24	Anthracene	0.976	0.968	0.8	97	-0.01
25 SURR	Fluoranthene-d10	5.280	5.307	-0.5	97	-0.02
26	Fluoranthene	1.528	1.572	-2.9	100	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	94	-0.02
28	Pyrene	1.056	1.092	-3.4	99	-0.02
29 S	Terphenyl-d14	0.730	0.768	-5.2	99	-0.02
30	Benzo(a)anthracene	1.259	1.258	0.1	93	-0.02
31	Chrysene	1.235	1.290	-4.5	97	-0.01
32	Bis(2-ethylhexyl)phthalate	1.605	1.044	35.0#	63	-0.02
33	Indeno(1,2,3-cd)pyrene	1.412	1.488	-5.4	96	-0.06
34 I	Perylene-d12	1.000	1.000	0.0	91	-0.03
35	Benzo(b)fluoranthene	1.194	1.265	-5.9	100	-0.03
36	Benzo(k)fluoranthene	1.162	1.270	-9.3	102	-0.03
37 C	Benzo(a)pyrene	1.122	1.144	-2.0	96	-0.03
38	Dibenzo(a,h)anthracene	1.116	1.191	-6.7	99	-0.07
39	Benzo(g,h,i)perylene	1.112	1.158	-4.1	97	-0.07

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE040419\
Data File : BE099293.D
Acq On : 4 Apr 2019 16:00
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 05 04:29:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 02 17:16:42 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			