

Data Path : Z:\HPCHEM1\BNA E\DATA\BE040518\
 Data File : BE095951.D
 Acq On : 5 Apr 2018 10:32
 Operator : SJ/JU
 Sample : J2148-03RE
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SB-2BRE

Manual Integrations
 APPROVED

Sohil
 4/6/2018 5:26:27 PM

Quant Time: Apr 06 12:42:19 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 12:41:11 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.42	152	652	0.40	ng	0.00
7) Naphthalene-d8	11.26	136	2576	0.40	ng	0.00
13) Acenaphthene-d10	14.99	164	1439	0.40	ng	-0.01
19) Phenanthrene-d10	17.70	188	3146	0.40	ng	-0.01
27) Chrysene-d12	21.79	240	3424	0.40	ng	-0.01
34) Perylene-d12	24.42	264	3451	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.91	112	1657	0.77	ng	0.00
5) Phenol-d6	7.56	99	2441	0.83	ng	0.00
8) Nitrobenzene-d5	9.59	82	2347	0.77	ng	0.00
11) 2-Methylnaphthalene-d10	12.79	152	3020	0.70	ng	-0.01
14) 2,4,6-Tribromophenol	16.46	330	636	0.77	ng	-0.01
15) 2-Fluorobiphenyl	13.63	172	4287	0.71	ng	-0.01
25) Fluoranthene-d10	19.68	212	24610	0.53	ng	0.00
29) Terphenyl-d14	20.24	244	4016	0.55	ng	0.00
Target Compounds						
9) Naphthalene	11.29	128	2131	0.072	ng	# 91
12) 2-Methylnaphthalene	12.86	142	120	0.024	ng	# 78
16) Acenaphthylene	14.73	152	359	0.044	ng	# 89
23) Phenanthrene	17.74	178	705	0.075	ng	# 93
24) Anthracene	17.84	178	222	0.024	ng	# 55
26) Fluoranthene	19.71	202	1273	0.097	ng	# 88
28) Pyrene	20.06	202	1370	0.098	ng	# 69
30) Benzo(a)anthracene	21.78	228	629m	0.053	ng	
31) Chrysene	21.82	228	763m	0.070	ng	
32) Bis(2-ethylhexyl)phthalate	21.64	149	19329	0.532	ng	# 96
33) Indeno(1,2,3-cd)pyrene	27.22	276	920	0.079	ng	# 78
39) Benzo(a,h,i)perylene	28.10	276	1118m	0.109	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE040518\
Data File : BE095951.D
Acq On : 5 Apr 2018 10:32
Operator : SJ/JU
Sample : J2148-03RE
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SB-2BRE

Quant Time: Apr 06 12:42:19 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 12:41:11 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

Sohil
4/6/2018 5:26:27 PM

