

Data Path : Z:\HPCHEM1\BNA E\Data\BE040518\
 Data File : BE095949.D
 Acq On : 5 Apr 2018 8:53
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 06 04:41:59 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 12:41:11 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	127	-0.02
2	1,4-Dioxane	0.400	0.390	2.5	130	0.00
3	n-Nitrosodimethylamine	0.400	0.380	5.0	117	0.00
4 S	2-Fluorophenol	0.400	0.391	2.3	119	-0.02
5 S	Phenol-d6	0.400	0.430	-7.5	131	0.00
6	bis(2-Chloroethyl)ether	0.400	0.428	-7.0	131	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	142	0.00
8 S	Nitrobenzene-d5	0.400	0.390	2.5	138	0.00
9	Naphthalene	0.400	0.417	-4.2	143	-0.01
10	Hexachlorobutadiene	0.400	0.365	8.8	125	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.415	-3.7	143	-0.01
12	2-Methylnaphthalene	0.400	0.427	-6.7	148	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	142	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.336	16.0	123	-0.01
15 S	2-Fluorobiphenyl	0.400	0.408	-2.0	141	-0.01
16	Acenaphthylene	0.400	0.407	-1.7	144	-0.01
17	Acenaphthene	0.400	0.411	-2.7	147	0.00
18	Fluorene	0.400	0.393	1.8	138	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	125	-0.01
20	4-Bromophenyl-phenylether	0.400	0.429	-7.2	132	0.00
21	Hexachlorobenzene	0.400	0.420	-5.0	129	0.00
22	Pentachlorophenol	0.400	0.494	-23.5	179	-0.01
23	Phenanthrene	0.400	0.420	-5.0	128	0.00
24	Anthracene	0.400	0.403	-0.8	125	-0.01
25 SURR	Fluoranthene-d10	0.400	0.364	9.0	112	-0.01
26	Fluoranthene	0.400	0.372	7.0	115	-0.01
27 I	Chrysene-d12	0.400	0.400	0.0	108	-0.01
28	Pvrene	0.400	0.446	-11.5	109	-0.01
29 S	Terphenyl-d14	0.400	0.402	-0.5	108	0.00
30	Benzo(a)anthracene	0.400	0.412	-3.0	109	-0.01
31	Chrysene	0.400	0.424	-6.0	113	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.375	6.3	108	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.430	-7.5	114	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	112	-0.02
35	Benzo(b)fluoranthene	0.400	0.416	-4.0	115	-0.01
36	Benzo(k)fluoranthene	0.400	0.417	-4.2	116	-0.02
37 C	Benzo(a)pyrene	0.400	0.417	-4.2	115	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.417	-4.2	113	-0.02
39	Benzo(a,h,i)perylene	0.400	0.417	-4.2	115	-0.02

Data Path : Z:\HPCHEM1\BNA E\Data\BE040518\
Data File : BE095949.D
Acq On : 5 Apr 2018 8:53
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 06 04:41:59 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE032318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 12:41:11 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			