

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040621\
 Data File : BE101211.D
 Acq On : 6 Apr 2021 11:40
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Quant Time: Apr 06 13:47:38 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE040621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 06 13:45:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.838	152	3647	0.40	ng	0.00
7) Naphthalene-d8	10.613	136	16586	0.40	ng	0.00
13) Acenaphthene-d10	14.443	164	11147	0.40	ng	0.00
19) Phenanthrene-d10	17.211	188	31260	0.40	ng	0.00
29) Chrysene-d12	21.366	240	24848	0.40	ng	# 0.01
36) Perylene-d12	23.840	264	24862	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.420	112	969	0.10	ng	0.00
5) Phenol-d6	6.986	99	1410	0.10	ng	0.00
8) Nitrobenzene-d5	8.976	82	1168	0.09	ng	0.00
11) 2-Methylnaphthalene-d10	12.196	152	3750	0.10	ng	0.00
14) 2,4,6-Tribromophenol	15.956	330	272	0.08	ng	0.00
15) 2-Fluorobiphenyl	13.075	172	4175	0.10	ng	0.00
27) Fluoranthene-d10	19.221	212	44749	0.08	ng	0.00
31) Terphenyl-d14	19.818	244	6474	0.12	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.370	88	726	0.11	ng	86
3) n-Nitrosodimethylamine	3.681	42	651	0.10	ng	# 78
6) bis(2-Chloroethyl)ether	7.251	93	1289	0.11	ng	95
9) Naphthalene	10.665	128	18138	0.10	ng	99
10) Hexachlorobutadiene	10.964	225	837	0.11	ng	99
12) 2-Methylnaphthalene	12.268	142	3123	0.10	ng	98
16) Acenaphthylene	14.170	152	3900	0.09	ng	99
17) Acenaphthene	14.515	154	3114	0.10	ng	98
18) Fluorene	15.502	166	4161	0.10	ng	98
20) 4,6-Dinitro-2-methylph...	15.578	198	219	0.05	ng	# 39
21) 4-Bromophenyl-phenylether	16.409	248	1441	0.09	ng	# 94
22) Hexachlorobenzene	16.515	284	1526	0.09	ng	98
23) Atrazine	16.666	200	945	0.08	ng	98
24) Pentachlorophenol	16.863	266	389	0.06	ng	93
25) Phenanthrene	17.241	178	9016	0.10	ng	99
26) Anthracene	17.332	178	6867	0.09	ng	98
28) Fluoranthene	19.249	202	9360	0.08	ng	100
30) Pyrene	19.606	202	9358	0.13	ng	100
32) Benzo(a)anthracene	21.342	228	6845	0.09	ng	98
33) Chrysene	21.402	228	8647	0.10	ng	96
34) Bis(2-ethylhexyl)phtha...	21.282	149	5782	0.06	ng	# 99
35) Indeno(1,2,3-cd)pyrene	26.445	276	6884	0.09	ng	99
37) Benzo(b)fluoranthene	23.077	252	7331	0.10	ng	# 96
38) Benzo(k)fluoranthene	23.131	252	7955	0.09	ng	# 90
39) Benzo(a)pyrene	23.722	252	5994	0.09	ng	# 91
40) Dibenzo(a,h)anthracene	26.480	278	5778	0.09	ng	94
41) Benzo(g,h,i)perylene	27.244	276	6931	0.10	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040621\
Data File : BE101211.D
Acq On : 6 Apr 2021 11:40
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampled :
SSTDICC0.1

Quant Time: Apr 06 13:47:38 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE040621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 06 13:45:08 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.838 min Scan# 432
 Delta R.T. 0.000 min
 Lab File: BE101211.D
 Acq: 6 Apr 2021 11:40




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:152 Resp: 3647
 Ion Ratio Lower Upper
 152 100
 150 152.8 116.9 175.3
 115 58.8 44.5 66.7



1,4-Dioxane
 Concen: 0.11 ng
 RT: 3.370 min Scan# 44
 Delta R.T. 0.000 min
 Lab File: BE101211.D
 Acq: 6 Apr 2021 11:40

Tgt Ion: 88 Resp: 726
 Ion Ratio Lower Upper
 88 100
 58 76.2 73.6 110.4
 43 29.8 27.6 41.4









n-Nitrosodimethylamine
 Concen: 0.10 ng
 RT: 3.681 min Scan# 71
 Delta R.T. 0.012 min
 Lab File: BE101211.D
 Acq: 6 Apr 2021 11:40




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1


Tgt Ion: 42 Resp: 651
 Ion Ratio Lower Upper
 42 100
 74 93.4 92.7 139.1
 44 0.0 10.7 16.1#



2-Fluorophenol
 Concen: 0.10 ng
 RT: 5.420 min Scan# 222
 Delta R.T. 0.000 min
 Lab File: BE101211.D
 Acq: 6 Apr 2021 11:40

Tgt Ion: 112 Resp: 969
 Ion Ratio Lower Upper
 112 100
 64 76.5 62.3 93.5
 63 161.6 117.3 175.9









Phenol-d6
 Concen: 0.10 ng
 RT: 6.986 min Scan# 358
 Delta R.T. 0.000 min
 Lab File: BE101211.D
 Acq: 6 Apr 2021 11:40




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	99	Resp:	1410
Ion	Ratio	Lower	Upper
99	100		
42	27.4	18.5	27.7
71	35.5	26.5	39.7




bis(2-Chloroethyl)ether
 Concen: 0.11 ng
 RT: 7.251 min Scan# 381
 Delta R.T. 0.000 min
 Lab File: BE101211.D
 Acq: 6 Apr 2021 11:40

Tgt Ion:	93	Resp:	1289
Ion	Ratio	Lower	Upper
93	100		
63	378.0	292.5	438.7
95	33.6	25.7	38.5





Abundance Scan 650 (10.615 min): BE101213.D\data.ms (-6457-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.613 min Scan# 650
Delta R.T. -0.002 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1


Tgt Ion:	136	Resp:	16586
Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.6	13.0
54	49.8	34.7	52.1
68	10.0	7.0	10.4



Abundance Scan 650 (10.613 min): BE101211.D\data.ms (-635-)





Abundance Scan 524 (8.978 min): BE101213.D\data.ms (-52178)




Nitrobenzene-d5
Concen: 0.09 ng
RT: 8.976 min Scan# 524
Delta R.T. -0.002 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40


Tgt Ion:	82	Resp:	1168
Ion	Ratio	Lower	Upper
82	100		
128	155.3	118.7	178.1
54	239.1	164.4	246.6



Abundance Scan 524 (8.976 min): BE101211.D\data.ms (-509-)






Abundance Scan 654 (10.667 min): BE101213.D\data.ms (-649) (-)




Naphthalene
Concen: 0.10 ng
RT: 10.665 min Scan# 654
Delta R.T. -0.002 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion:	128	Resp:	18138
Ion	Ratio	Lower	Upper
128	100		
129	3.2	2.2	3.4
127	3.6	2.6	3.8







Abundance Scan 677 (10.965 min): BE101213.D\data.ms (-672) (-)



Hexachlorobutadiene
Concen: 0.11 ng
RT: 10.964 min Scan# 677
Delta R.T. -0.002 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40

Tgt Ion:	225	Resp:	837
Ion	Ratio	Lower	Upper
225	100		
223	63.6	50.4	75.6
227	62.2	50.1	75.1









2-Methylnaphthalene-d10
 Concen: 0.10 ng
 RT: 12.196 min Scan# 763
 Delta R.T. -0.000 min
 Lab File: BE101211.D
 Acq: 6 Apr 2021 11:40




Instrument :
 BNA_E
 ClientSampled :
 SSTDICC0.1


Tgt Ion:152 Resp: 3750
 Ion Ratio Lower Upper
 152 100
 151 21.3 16.5 24.7



2-Methylnaphthalene
 Concen: 0.10 ng
 RT: 12.268 min Scan# 768
 Delta R.T. -0.000 min
 Lab File: BE101211.D
 Acq: 6 Apr 2021 11:40

Tgt Ion:142 Resp: 3123
 Ion Ratio Lower Upper
 142 100
 141 89.1 73.0 109.6
 115 34.2 27.8 41.8









Acenaphthene-d10
 Concen: 0.40 ng
 RT: 14.443 min Scan# 919
 Delta R.T. -0.000 min
 Lab File: BE101211.D
 Acq: 6 Apr 2021 11:40




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:	164	Resp:	11147
Ion	Ratio	Lower	Upper
164	100		
162	106.1	85.3	127.9
160	50.0	39.1	58.7



2,4,6-Tribromophenol
 Concen: 0.08 ng
 RT: 15.956 min Scan# 1016
 Delta R.T. 0.000 min
 Lab File: BE101211.D
 Acq: 6 Apr 2021 11:40

Tgt Ion:	330	Resp:	272
Ion	Ratio	Lower	Upper
330	100		
332	87.5	77.8	116.6
141	44.5	30.4	45.6









2-Fluorobiphenyl
Concen: 0.10 ng
RT: 13.075 min Scan# 824
Delta R.T. -0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1


Tgt Ion:	172	Resp:	4175
Ion	Ratio	Lower	Upper
172	100		
171	36.9	28.7	43.1
170	25.8	19.3	28.9



Acenaphthylene
Concen: 0.09 ng
RT: 14.170 min Scan# 900
Delta R.T. -0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40

Tgt Ion:	152	Resp:	3900
Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.4	23.2
153	13.6	10.0	15.0









Acenaphthene
Concen: 0.10 ng
RT: 14.515 min Scan# 924
Delta R.T. -0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion:154 Resp: 3114
Ion Ratio Lower Upper
154 100
153 108.8 84.9 127.3
152 54.5 42.7 64.1



Fluorene
Concen: 0.10 ng
RT: 15.502 min Scan# 986
Delta R.T. 0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40


Tgt Ion:166 Resp: 4161
Ion Ratio Lower Upper
166 100
165 96.4 79.0 118.4
167 13.1 11.0 16.4






Abundance Scan 1099 (17.211 min): BE101213.D\data.ms (-1084) (-)

Phenanthrene-d10
Concen: 0.40 ng
RT: 17.211 min Scan# 1099
Delta R.T. 0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1






Tgt Ion:	188	Resp:	31260
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	6.5	4.6	6.8



Abundance Scan 991 (15.578 min): BE101213.D\data.ms (-984) (-)

4,6-Dinitro-2-methylphenol
Concen: 0.05 ng
RT: 15.578 min Scan# 991
Delta R.T. 0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40


Tgt Ion:	198	Resp:	219
Ion	Ratio	Lower	Upper
198	100		
51	131.7	55.9	83.9#
105	65.6	33.6	50.4#






Abundance Scan 1046 (16.409 min): BE101213.D\data.ms (-1042) (-)

4-Bromophenyl-phenylether
Concen: 0.09 ng
RT: 16.409 min Scan# 1046
Delta R.T. 0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1




Tgt Ion:	248	Resp:	1441
Ion	Ratio	Lower	Upper
248	100		
250	101.4	84.5	126.7
141	31.3	19.8	29.8#






Abundance Scan 1053 (16.515 min): BE101213.D\data.ms (-1049) (-)

Hexachlorobenzene
Concen: 0.09 ng
RT: 16.515 min Scan# 1053
Delta R.T. 0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40




Tgt Ion:	284	Resp:	1526
Ion	Ratio	Lower	Upper
284	100		
142	33.9	27.4	41.0
249	34.7	26.5	39.7






Abundance Scan 1063 (16.666 min): BE101213.D\data.ms (-1030) (-)

Atrazine
Concen: 0.08 ng
RT: 16.666 min Scan# 1063
Delta R.T. 0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1




Tgt Ion:	200	Resp:	945
Ion	Ratio	Lower	Upper
200	100		
173	29.5	21.9	32.9
215	49.0	38.6	57.8







Abundance Scan 1076 (16.863 min): BE101213.D\data.ms (-1024) (-)

Pentachlorophenol
Concen: 0.06 ng
RT: 16.863 min Scan# 1076
Delta R.T. 0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40



Tgt Ion:	266	Resp:	389
Ion	Ratio	Lower	Upper
266	100		
264	60.4	49.1	73.7
268	53.7	50.6	76.0









Phenanthrene
Concen: 0.10 ng
RT: 17.241 min Scan# 1101
Delta R.T. 0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion:	178	Resp:	9016
Ion	Ratio	Lower	Upper
178	100		
176	19.6	15.0	22.6
179	15.7	12.2	18.4




Anthracene
Concen: 0.09 ng
RT: 17.332 min Scan# 1107
Delta R.T. 0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40

Tgt Ion:	178	Resp:	6867
Ion	Ratio	Lower	Upper
178	100		
176	18.8	14.4	21.6
179	15.4	11.8	17.8





Abundance Scan 1359 (19.220 min): BE101213.D\data.ms (-1327) (-)




Fluoranthene-d10
Concen: 0.08 ng
RT: 19.221 min Scan# 1359
Delta R.T. 0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1


Tgt Ion:	212	Resp:	44749
Ion	Ratio	Lower	Upper
212	100		
106	3.2	2.6	4.0
104	1.8	1.4	2.2



Abundance Scan 1359 (19.221 min): BE101211.D\data.ms (-1324) (-)





Abundance Scan 1364 (19.249 min): BE101213.D\data.ms (-1328) (-)




Fluoranthene
Concen: 0.08 ng
RT: 19.249 min Scan# 1364
Delta R.T. 0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40

Tgt Ion:	202	Resp:	9360
Ion	Ratio	Lower	Upper
202	100		
101	11.5	9.1	13.7
203	17.1	13.6	20.4




Abundance Scan 1364 (19.249 min): BE101211.D\data.ms (-1329) (-)






Abundance Scan 1603 (21.354 min): BE101213.D\data.ms (-1529) (-)

Chrysene-d12
Concen: 0.40 ng
RT: 21.366 min Scan# 1604
Delta R.T. 0.012 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1




Tgt Ion:	240	Resp:	24848
Ion	Ratio	Lower	Upper
240	100		
120	4.6	10.2	15.4#
236	25.3	21.7	32.5






Abundance Scan 1426 (19.605 min): BE101213.D\data.ms (-1439) (-)

Pyrene
Concen: 0.13 ng
RT: 19.606 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40




Tgt Ion:	202	Resp:	9358
Ion	Ratio	Lower	Upper
202	100		
200	20.8	16.6	24.8
203	17.0	13.8	20.8






Abundance Scan 1463 (19.818 min): BE101213.D\data.ms (-1434) (-)

Terphenyl-d14
Concen: 0.12 ng
RT: 19.818 min Scan# 1463
Delta R.T. 0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1




Tgt Ion:	244	Resp:	6474
Ion	Ratio	Lower	Upper
244	100		
212	30.4	21.9	32.9
122	9.7	7.3	10.9







Abundance Scan 1602 (21.342 min): BE101213.D\data.ms (-1594) (-)

Benzo(a)anthracene
Concen: 0.09 ng
RT: 21.342 min Scan# 1602
Delta R.T. 0.000 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40



Tgt Ion:	228	Resp:	6845
Ion	Ratio	Lower	Upper
228	100		
226	28.0	21.4	32.2
229	20.7	15.7	23.5









Chrysene
 Concen: 0.10 ng
 RT: 21.402 min Scan# 1607
 Delta R.T. 0.012 min
 Lab File: BE101211.D
 Acq: 6 Apr 2021 11:40




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	228	Resp:	8647
Ion	Ratio	Lower	Upper
228	100		
226	28.8	24.8	37.2
229	20.3	15.2	22.8




Bis(2-ethylhexyl)phthalate
 Concen: 0.06 ng
 RT: 21.282 min Scan# 1597
 Delta R.T. 0.000 min
 Lab File: BE101211.D
 Acq: 6 Apr 2021 11:40

Tgt Ion:	149	Resp:	5782
Ion	Ratio	Lower	Upper
149	100		
167	7.3	5.5	8.3
279	1.6	1.0	1.4#






Abundance Scan 2893 (26.448 min): BE101213.D\data.ms (-2865) (-)




Indeno(1,2,3-cd)pyrene
Concen: 0.09 ng
RT: 26.445 min Scan# 2892
Delta R.T. -0.003 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion:	276	Resp:	6884
Ion	Ratio	Lower	Upper
276	100		
138	20.6	17.4	26.0
277	24.7	19.8	29.8






Abundance Scan 2165 (23.838 min): BE101213.D\data.ms (-2123) (-)



Perylene-d12
Concen: 0.40 ng
RT: 23.840 min Scan# 2166
Delta R.T. 0.002 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40


Tgt Ion:	264	Resp:	24862
Ion	Ratio	Lower	Upper
264	100		
260	25.0	19.3	28.9
265	27.1	22.4	33.6





Abundance Scan 1951 (23.075 min): BE101213.D\data.ms (-1937) (-)

Benzo(b)fluoranthene
Concen: 0.10 ng
RT: 23.077 min Scan# 1952
Delta R.T. 0.002 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40


Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1



Tgt Ion:	252	Resp:	7331
Ion	Ratio	Lower	Upper
252	100		
253	23.9	18.1	27.1
125	12.5	7.6	11.4#





Abundance Scan 1952 (23.077 min): BE101211.D\data.ms (-1895) (-)




Abundance Scan 1965 (23.125 min): BE101213.D\data.ms (-1938) (-)

Benzo(k)fluoranthene
Concen: 0.09 ng
RT: 23.131 min Scan# 1967
Delta R.T. 0.006 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40


Tgt Ion:	252	Resp:	7955
Ion	Ratio	Lower	Upper
252	100		
253	25.5	17.8	26.6
125	16.0	7.6	11.4#



Abundance Scan 1967 (23.131 min): BE101211.D\data.ms (-1881) (-)






Abundance Scan 2133 (23.724 min): BE101213.D\data.ms (-2#89) (-)




Benzo(a)pyrene
Concen: 0.09 ng
RT: 23.722 min Scan# 2133
Delta R.T. -0.002 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion:252 Resp: 5994
Ion Ratio Lower Upper
252 100
253 26.4 18.0 27.0
125 14.7 8.0 12.0#






Abundance Scan 2900 (26.473 min): BE101213.D\data.ms (-2#90) (-)




Dibenzo(a,h)anthracene
Concen: 0.09 ng
RT: 26.480 min Scan# 2902
Delta R.T. 0.007 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40

Tgt Ion:278 Resp: 5778
Ion Ratio Lower Upper
278 100
139 17.1 12.2 18.2
279 27.6 19.3 28.9





Abundance Scan 3117 (27.247 min): BE101213.D\data.ms (-3091) (-)



Benzo(g,h,i)perylene
Concen: 0.10 ng
RT: 27.244 min Scan# 3116
Delta R.T. -0.003 min
Lab File: BE101211.D
Acq: 6 Apr 2021 11:40

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.3	19.4	29.0
138	20.8	15.2	22.8



Abundance Scan 3116 (27.244 min): BE101211.D\data.ms (-3061) (-)

