

Data Path : Z:\HPCHEM1\BNA E\DATA\BE040816\
 Data File : BE091614.D
 Acq On : 8 Apr 2016 16:28
 Operator : UM/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE040816

Quant Time: Apr 08 17:03:32 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 08 16:48:14 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	-0.02
2	1,4-Dioxane	0.640	0.684	-6.9	104	0.00
3	n-Nitrosodimethylamine	0.767	0.687	10.4	95	0.00
4 S	2-Fluorophenol	1.193	1.111	6.9	97	0.00
5 S	Phenol-d6	1.591	1.439	9.6	98	0.00
6	bis(2-Chloroethyl)ether	1.379	1.298	5.9	98	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.290	0.245	15.5	96	0.00
9	Nitrobenzene	0.305	0.256	16.1	97	0.00
10	Naphthalene	1.031	1.013	1.7	101	0.00
11	Hexachlorobutadiene	0.152	0.148	2.6	100	0.00
12	2-Methylnaphthalene	0.658	0.642	2.4	102	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.138	0.106	23.2	95	0.00
15 S	2-Fluorobiphenyl	1.390	1.381	0.6	100	0.00
16	Acenaphthylene	1.941	1.829	5.8	115	0.00
17	Acenaphthene	1.227	1.204	1.9	101	0.00
18	Fluorene	1.533	1.482	3.3	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
20	4-Bromophenyl-phenylether	0.176	0.167	5.1	97	0.00
21	Hexachlorobenzene	0.183	0.183	0.0	99	0.00
22	Pentachlorophenol	0.068	0.046	32.4#	97	0.00
23	Phenanthrene	1.141	1.134	0.6	99	0.00
24	Anthracene	1.011	0.951	5.9	99	0.00
25	Fluoranthene	1.208	1.114	7.8	98	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
27	Benzydine	0.329	0.195	40.7#	87	0.00
28	Pvrene	1.000	0.883	11.7	97	0.00
29 S	Terphenyl-d14	0.624	0.589	5.6	97	0.00
30	Benzo(a)anthracene	1.119	1.042	6.9	98	0.00
31	3,3'-Dichlorobenzidine	0.599	0.430	28.2#	87	0.00
32	Chrysene	1.015	1.007	0.8	102	-0.02
33	Bis(2-ethylhexyl)phthalate	0.494	0.279	43.5#	88	0.00
34	Indeno(1,2,3-cd)pyrene	1.132	1.073	5.2	99	-0.01
35 I	Perylene-d12	1.000	1.000	0.0	100	0.00
36	Benzo(b)fluoranthene	1.161	1.144	1.5	103	0.00
37	Benzo(k)fluoranthene	1.145	1.062	7.2	97	0.00
38 C	Benzo(a)pyrene	1.079	1.013	6.1	99	0.00
39	Dibenzo(a,h)anthracene	1.062	1.001	5.7	99	0.00
40	Benzo(a,h,i)perylene	1.094	1.042	4.8	98	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE040816\
Data File : BE091614.D
Acq On : 8 Apr 2016 16:28
Operator : UM/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE040816

Quant Time: Apr 08 17:03:32 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 08 16:48:14 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				