

Data Path : Z:\HPCHEM1\BNA E\DATA\BE040918\
 Data File : BE095962.D
 Acq On : 9 Apr 2018 11:02
 Operator : SJ/JU
 Sample : PB107961BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Apr 10 03:47:30 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 10 03:47:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.38	152	1900	0.40	ng/ul	0.00
2) Naphthalene-d8	11.21	136	4969	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.97	164	2794	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.68	188	6997	0.40	ng/ul	0.00
16) Chrysene-d12	21.77	240	8900	0.40	ng/ul	0.00
20) Perylene-d12	24.40	264	7144	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.77	152	3419	0.37	ng/ul	0.00
14) Fluoranthene-d10	19.67	212	9940	0.34	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE040918\
Data File : BE095962.D
Acq On : 9 Apr 2018 11:02
Operator : SJ/JU
Sample : PB107961BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Apr 10 03:47:30 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE040618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 10 03:47:21 2018
Response via : Initial Calibration

