

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040921\
 Data File : BE101266.D
 Acq On : 9 Apr 2021 10:02
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 09 11:10:19 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE040621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 07 10:22:45 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	66	0.00
2	1,4-Dioxane	0.400	0.398	0.5	63	0.00
3	n-Nitrosodimethylamine	0.400	0.488	-22.0	82	0.00
4 S	2-Fluorophenol	0.400	0.475	-18.7	79	0.00
5 S	Phenol-d6	0.400	0.489	-22.2	84	0.00
6	bis(2-Chloroethyl)ether	0.400	0.413	-3.2	67	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	71	0.00
8 S	Nitrobenzene-d5	0.400	0.459	-14.8	86	0.00
9	Naphthalene	0.400	0.402	-0.5	71	0.00
10	Hexachlorobutadiene	0.400	0.400	0.0	70	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.413	-3.2	75	0.00
12	2-Methylnaphthalene	0.400	0.408	-2.0	76	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	84	0.00
14 S	2,4,6-Tribromophenol	0.400	0.428	-7.0	99	0.00
15 S	2-Fluorobiphenyl	0.400	0.377	5.8	74	0.00
16	Acenaphthylene	0.400	0.405	-1.3	89	0.00
17	Acenaphthene	0.400	0.379	5.3	80	0.00
18	Fluorene	0.400	0.398	0.5	87	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	94	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.219	45.3#	43	0.00
21	4-Bromophenyl-phenylether	0.400	0.401	-0.3	92	-0.01
22	Hexachlorobenzene	0.400	0.383	4.3	84	0.00
23	Atrazine	0.400	0.453	-13.2	117	0.00
24	Pentachlorophenol	0.400	0.132	67.0#	43	0.00
25	Phenanthrene	0.400	0.401	-0.3	93	0.00
26	Anthracene	0.400	0.409	-2.2	103	0.00
27 SURR	Fluoranthene-d10	0.400	0.414	-3.5	118	0.00
28	Fluoranthene	0.400	0.409	-2.2	117	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	91	0.00
30	Pyrene	0.400	0.474	-18.5	121	0.00
31 S	Terphenyl-d14	0.400	0.465	-16.3	118	0.00
32	Benzo(a)anthracene	0.400	0.426	-6.5	101	0.00
33	Chrysene	0.400	0.404	-1.0	90	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.621	-55.2#	145	-0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.495	-23.7	102	0.00
36 I	Perylene-d12	0.400	0.400	0.0	96	0.00
37	Benzo(b)fluoranthene	0.400	0.373	6.8	92	0.00
38	Benzo(k)fluoranthene	0.400	0.387	3.3	97	0.00
39 C	Benzo(a)pyrene	0.400	0.405	-1.3	103	0.00
40	Dibenzo(a,h)anthracene	0.400	0.417	-4.2	101	0.00
41	Benzo(g,h,i)perylene	0.400	0.400	0.0	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_E

LabSampleId :

SSTDCCC0.4