

Data Path : Z:\HPCHEM1\BNA E\Data\BE041116\
 Data File : BE091620.D
 Acq On : 11 Apr 2016 10:02
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 12 14:53:36 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 08 16:48:14 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	129	0.00
2	1,4-Dioxane	0.640	0.648	-1.3	129	0.00
3	n-Nitrosodimethylamine	0.767	0.724	5.6	131	-0.02
4 S	2-Fluorophenol	1.193	1.079	9.6	124	0.00
5 S	Phenol-d6	1.591	1.437	9.7	128	0.00
6	bis(2-Chloroethyl)ether	1.379	1.328	3.7	132	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	132	0.00
8 S	Nitrobenzene-d5	0.290	0.243	16.2	125	0.00
9	Nitrobenzene	0.305	0.247	19.0	123	0.00
10	Naphthalene	1.031	1.010	2.0	131	0.00
11	Hexachlorobutadiene	0.152	0.148	2.6	130	0.00
12	2-Methylnaphthalene	0.658	0.634	3.6	131	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	128	0.00
14 S	2,4,6-Tribromophenol	0.138	0.101	26.8#	117	0.00
15 S	2-Fluorobiphenyl	1.390	1.387	0.2	130	0.00
16	Acenaphthylene	1.941	1.777	8.4	145	0.00
17	Acenaphthene	1.227	1.255	-2.3	136	0.00
18	Fluorene	1.533	1.463	4.6	126	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	125	0.00
20	4-Bromophenyl-phenylether	0.176	0.166	5.7	122	0.00
21	Hexachlorobenzene	0.183	0.182	0.5	126	0.00
22	Pentachlorophenol	0.068	0.052	23.5	139	0.00
23	Phenanthrene	1.141	1.133	0.7	125	0.00
24	Anthracene	1.011	0.942	6.8	124	0.00
25	Fluoranthene	1.208	1.100	8.9	123	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	134	0.00
27	Benzydine	0.329	0.224	31.9#	133	0.00
28	Pvrene	1.000	0.867	13.3	126	0.00
29 S	Terphenyl-d14	0.624	0.563	9.8	123	0.00
30	Benzo(a)anthracene	1.119	1.093	2.3	136	0.00
31	3,3'-Dichlorobenzidine	0.599	0.412	31.2#	111	0.00
32	Chrysene	1.015	0.996	1.9	134	-0.02
33	Bis(2-ethylhexyl)phthalate	0.494	0.238	51.8#	100	0.00
34	Indeno(1,2,3-cd)pyrene	1.132	1.000	11.7	123	-0.01
35 I	Perylene-d12	1.000	1.000	0.0	129	0.00
36	Benzo(b)fluoranthene	1.161	1.170	-0.8	136	0.00
37	Benzo(k)fluoranthene	1.145	0.994	13.2	117	0.00
38 C	Benzo(a)pyrene	1.079	0.978	9.4	124	0.00
39	Dibenzo(a,h)anthracene	1.062	0.966	9.0	123	-0.01
40	Benzo(a,h,i)perylene	1.094	1.008	7.9	123	-0.01

Data Path : Z:\HPCHEM1\BNA E\Data\BE041116\
Data File : BE091620.D
Acq On : 11 Apr 2016 10:02
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 12 14:53:36 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 08 16:48:14 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				