

Data Path : Z:\HPCHEM1\BNA E\Data\BE041116\
 Data File : BE091620.D
 Acq On : 11 Apr 2016 10:02
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 12 14:53:36 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 08 16:48:14 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	129	0.00
2	1,4-Dioxane	0.400	0.392	2.0	129	0.00
3	n-Nitrosodimethylamine	0.400	0.378	5.5	131	-0.02
4 S	2-Fluorophenol	0.400	0.362	9.5	124	0.00
5 S	Phenol-d6	0.400	0.361	9.8	128	0.00
6	bis(2-Chloroethyl)ether	0.400	0.385	3.8	132	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	132	0.00
8 S	Nitrobenzene-d5	0.400	0.335	16.3	125	0.00
9	Nitrobenzene	0.400	0.435	-8.7	123	0.00
10	Naphthalene	0.400	0.392	2.0	131	0.00
11	Hexachlorobutadiene	0.400	0.391	2.3	130	0.00
12	2-Methylnaphthalene	0.400	0.385	3.8	131	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	128	0.00
14 S	2,4,6-Tribromophenol	0.400	0.456	-14.0	117	0.00
15 S	2-Fluorobiphenyl	0.400	0.399	0.3	130	0.00
16	Acenaphthylene	0.400	0.366	8.5	145	0.00
17	Acenaphthene	0.400	0.409	-2.2	136	0.00
18	Fluorene	0.400	0.382	4.5	126	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	125	0.00
20	4-Bromophenyl-phenylether	0.400	0.376	6.0	122	0.00
21	Hexachlorobenzene	0.400	0.398	0.5	126	0.00
22	Pentachlorophenol	0.400	0.437	-9.2	139	0.00
23	Phenanthrene	0.400	0.397	0.8	125	0.00
24	Anthracene	0.400	0.373	6.8	124	0.00
25	Fluoranthene	0.400	0.364	9.0	123	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	134	0.00
27	Benzydine	0.400	0.502	-25.5#	133	0.00
28	Pvrene	0.400	0.347	13.3	126	0.00
29 S	Terphenyl-d14	0.400	0.361	9.8	123	0.00
30	Benzo(a)anthracene	0.400	0.391	2.3	136	0.00
31	3,3'-Dichlorobenzidine	0.400	0.355	11.3	111	0.00
32	Chrysene	0.400	0.393	1.8	134	-0.02
33	Bis(2-ethylhexyl)phthalate	0.400	0.373	6.8	100	0.00
34	Indeno(1,2,3-cd)pyrene	0.400	0.353	11.8	123	-0.01
35 I	Perylene-d12	5.000	5.000	0.0	129	0.00
36	Benzo(b)fluoranthene	0.400	0.403	-0.8	136	0.00
37	Benzo(k)fluoranthene	0.400	0.347	13.3	117	0.00
38 C	Benzo(a)pyrene	0.400	0.363	9.3	124	0.00
39	Dibenzo(a,h)anthracene	0.400	0.364	9.0	123	-0.01
40	Benzo(a,h,i)perylene	0.400	0.368	8.0	123	-0.01

Data Path : Z:\HPCHEM1\BNA E\Data\BE041116\
Data File : BE091620.D
Acq On : 11 Apr 2016 10:02
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 12 14:53:36 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 08 16:48:14 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)
----------	--------	-------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0