

Data Path : Z:\HPCHEM1\BNA E\Data\BE041116\
 Data File : BE091643.D
 Acq On : 12 Apr 2016 00:15
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 12 15:50:52 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 08 16:48:14 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	116	-0.02
2	1,4-Dioxane	0.400	0.351	12.3	105	0.00
3	n-Nitrosodimethylamine	0.400	0.386	3.5	120	-0.02
4 S	2-Fluorophenol	0.400	0.396	1.0	121	0.00
5 S	Phenol-d6	0.400	0.398	0.5	126	0.00
6	bis(2-Chloroethyl)ether	0.400	0.391	2.3	119	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	120	-0.02
8 S	Nitrobenzene-d5	0.400	0.395	1.3	134	0.00
9	Nitrobenzene	0.400	0.483	-20.7	132	0.00
10	Naphthalene	0.400	0.393	1.8	120	0.00
11	Hexachlorobutadiene	0.400	0.398	0.5	121	0.00
12	2-Methylnaphthalene	0.400	0.400	0.0	124	-0.01
13 I	Acenaphthene-d10	5.000	5.000	0.0	122	0.00
14 S	2,4,6-Tribromophenol	0.400	0.532	-33.0#	150	-0.01
15 S	2-Fluorobiphenyl	0.400	0.393	1.8	122	0.00
16	Acenaphthylene	0.400	0.400	0.0	151	0.00
17	Acenaphthene	0.400	0.412	-3.0	130	-0.01
18	Fluorene	0.400	0.394	1.5	124	-0.01
19 I	Phenanthrene-d10	5.000	5.000	0.0	129	-0.01
20	4-Bromophenyl-phenylether	0.400	0.358	10.5	120	0.00
21	Hexachlorobenzene	0.400	0.372	7.0	121	0.00
22	Pentachlorophenol	0.400	0.455	-13.7	156	0.00
23	Phenanthrene	0.400	0.369	7.8	120	0.00
24	Anthracene	0.400	0.375	6.3	129	0.00
25	Fluoranthene	0.400	0.371	7.3	129	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	123	0.00
27	Benzydine	0.400	0.552	-38.0#	159	-0.02
28	Pvrene	0.400	0.391	2.3	131	-0.02
29 S	Terphenyl-d14	0.400	0.439	-9.7	139	0.00
30	Benzo(a)anthracene	0.400	0.430	-7.5	139	0.00
31	3,3'-Dichlorobenzidine	0.400	0.439	-9.7	147	0.00
32	Chrysene	0.400	0.455	-13.7	144	-0.02
33	Bis(2-ethylhexyl)phthalate	0.400	0.438	-9.5	148	0.00
34	Indeno(1,2,3-cd)pyrene	0.400	0.396	1.0	127	-0.01
35 I	Perylene-d12	5.000	5.000	0.0	124	-0.01
36	Benzo(b)fluoranthene	0.400	0.393	1.8	128	-0.01
37	Benzo(k)fluoranthene	0.400	0.442	-10.5	144	0.00
38 C	Benzo(a)pyrene	0.400	0.399	0.3	132	0.00
39	Dibenzo(a,h)anthracene	0.400	0.385	3.8	126	-0.01
40	Benzo(a,h,i)perylene	0.400	0.391	2.3	126	-0.01

Data Path : Z:\HPCHEM1\BNA E\Data\BE041116\
Data File : BE091643.D
Acq On : 12 Apr 2016 00:15
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 12 15:50:52 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 08 16:48:14 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				