

Data Path : Z:\HPCHEM1\BNA E\DATA\BE041216\
 Data File : BE091649.D
 Acq On : 12 Apr 2016 13:58
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 13 01:21:31 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 08 16:48:14 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	124	-0.02
2	1,4-Dioxane	0.400	0.301	24.8	100	0.00
3	n-Nitrosodimethylamine	0.400	0.388	3.0	129	-0.02
4 S	2-Fluorophenol	0.400	0.364	9.0	119	-0.02
5 S	Phenol-d6	0.400	0.362	9.5	123	0.00
6	bis(2-Chloroethyl)ether	0.400	0.384	4.0	126	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	128	-0.02
8 S	Nitrobenzene-d5	0.400	0.356	11.0	129	0.00
9	Nitrobenzene	0.400	0.455	-13.7	128	-0.01
10	Naphthalene	0.400	0.390	2.5	127	0.00
11	Hexachlorobutadiene	0.400	0.391	2.3	126	0.00
12	2-Methylnaphthalene	0.400	0.388	3.0	128	-0.01
13 I	Acenaphthene-d10	5.000	5.000	0.0	129	0.00
14 S	2,4,6-Tribromophenol	0.400	0.496	-24.0	139	-0.01
15 S	2-Fluorobiphenyl	0.400	0.392	2.0	128	0.00
16	Acenaphthylene	0.400	0.375	6.3	149	0.00
17	Acenaphthene	0.400	0.413	-3.2	138	-0.01
18	Fluorene	0.400	0.394	1.5	131	-0.01
19 I	Phenanthrene-d10	5.000	5.000	0.0	131	0.00
20	4-Bromophenyl-phenylether	0.400	0.395	1.3	134	0.00
21	Hexachlorobenzene	0.400	0.396	1.0	130	0.00
22	Pentachlorophenol	0.400	0.441	-10.2	148	0.00
23	Phenanthrene	0.400	0.401	-0.3	132	0.00
24	Anthracene	0.400	0.393	1.8	136	0.00
25	Fluoranthene	0.400	0.362	9.5	127	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	129	0.00
27	Benzydine	0.400	0.510	-27.5#	134	-0.02
28	Pvrene	0.400	0.367	8.3	129	0.00
29 S	Terphenyl-d14	0.400	0.397	0.8	131	0.00
30	Benzo(a)anthracene	0.400	0.388	3.0	131	0.00
31	3,3'-Dichlorobenzidine	0.400	0.357	10.8	108	0.00
32	Chrysene	0.400	0.409	-2.2	135	-0.02
33	Bis(2-ethylhexyl)phthalate	0.400	0.369	7.8	94	0.00
34	Indeno(1,2,3-cd)pyrene	0.400	0.350	12.5	118	0.00
35 I	Perylene-d12	5.000	5.000	0.0	122	0.00
36	Benzo(b)fluoranthene	0.400	0.398	0.5	127	0.00
37	Benzo(k)fluoranthene	0.400	0.381	4.8	121	0.00
38 C	Benzo(a)pyrene	0.400	0.363	9.3	118	0.00
39	Dibenzo(a,h)anthracene	0.400	0.378	5.5	121	0.00
40	Benzo(a,h,i)perylene	0.400	0.367	8.3	115	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE041216\
Data File : BE091649.D
Acq On : 12 Apr 2016 13:58
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 13 01:21:31 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 08 16:48:14 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				