

Data Path : Z:\HPCHEM1\BNA E\Data\BE041216\
 Data File : BE091649.D
 Acq On : 12 Apr 2016 13:58
 Operator : UM/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Apr 13 01:21:31 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 08 16:48:14 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	124	-0.02
2	1,4-Dioxane	0.640	0.526	17.8	100	0.00
3	n-Nitrosodimethylamine	0.767	0.744	3.0	129	-0.02
4 S	2-Fluorophenol	1.193	1.086	9.0	119	-0.02
5 S	Phenol-d6	1.591	1.439	9.6	123	0.00
6	bis(2-Chloroethyl)ether	1.379	1.324	4.0	126	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	128	-0.02
8 S	Nitrobenzene-d5	0.290	0.259	10.7	129	0.00
9	Nitrobenzene	0.305	0.266	12.8	128	-0.01
10	Naphthalene	1.031	1.005	2.5	127	0.00
11	Hexachlorobutadiene	0.152	0.148	2.6	126	0.00
12	2-Methylnaphthalene	0.658	0.638	3.0	128	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	129	0.00
14 S	2,4,6-Tribromophenol	0.138	0.120	13.0	139	-0.01
15 S	2-Fluorobiphenyl	1.390	1.362	2.0	128	0.00
16	Acenaphthylene	1.941	1.817	6.4	149	0.00
17	Acenaphthene	1.227	1.268	-3.3	138	-0.01
18	Fluorene	1.533	1.511	1.4	131	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	131	0.00
20	4-Bromophenyl-phenylether	0.176	0.174	1.1	134	0.00
21	Hexachlorobenzene	0.183	0.181	1.1	130	0.00
22	Pentachlorophenol	0.068	0.053	22.1	148	0.00
23	Phenanthrene	1.141	1.144	-0.3	132	0.00
24	Anthracene	1.011	0.993	1.8	136	0.00
25	Fluoranthene	1.208	1.092	9.6	127	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	129	0.00
27	Benzydine	0.329	0.234	28.9#	134	-0.02
28	Pvrene	1.000	0.918	8.2	129	0.00
29 S	Terphenyl-d14	0.624	0.619	0.8	131	0.00
30	Benzo(a)anthracene	1.119	1.086	2.9	131	0.00
31	3,3'-Dichlorobenzidine	0.599	0.416	30.6#	108	0.00
32	Chrysene	1.015	1.037	-2.2	135	-0.02
33	Bis(2-ethylhexyl)phthalate	0.494	0.232	53.0#	94	0.00
34	Indeno(1,2,3-cd)pyrene	1.132	0.992	12.4	118	0.00
35 I	Perylene-d12	1.000	1.000	0.0	122	0.00
36	Benzo(b)fluoranthene	1.161	1.154	0.6	127	0.00
37	Benzo(k)fluoranthene	1.145	1.090	4.8	121	0.00
38 C	Benzo(a)pyrene	1.079	0.979	9.3	118	0.00
39	Dibenzo(a,h)anthracene	1.062	1.003	5.6	121	0.00
40	Benzo(a,h,i)perylene	1.094	1.005	8.1	115	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE041216\
Data File : BE091649.D
Acq On : 12 Apr 2016 13:58
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 13 01:21:31 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE040816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 08 16:48:14 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				